



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005  
& ANSI/NCSL Z540-1-1994

MICRO PRECISION CALIBRATION  
 3901 N. Highway 301, Suite 105  
 Tampa, Florida 33619  
 Anthony Barbera Phone: (813)621-6495

CALIBRATION

Valid To: July 31, 2011

Certificate Number: 935.09

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations<sup>1</sup>:

I. Dimensional

Parameter/Equipment	Range	Best Uncertainty <sup>2, 6</sup> ( $\pm$ )	Comments
Calipers & Height Gages <sup>3</sup>	(0.10 to 36) in (36 to 48) in	(56 $\mu$ in + 0.6L) $\mu$ in (1167 + 0.6L) $\mu$ in	Mitutoyo gage blocks
Gage Blocks	(0 to 12) in	(5 + 0.6L) $\mu$ in	Federal gage block comparator and master gage blocks
Surface Plates <sup>3</sup>  Surface Repeatability	12 in to 30 ft	40 $\mu$ in	Repeat-O-Meter (only valid in connection with flatness calibration)
Cylindrical Plug Gages	(0 to 12) in	(12 + 0.34L) $\mu$ in	Sip measuring machine

Parameter/Equipment	Range	Best Uncertainty <sup>2, 6</sup> ( $\pm$ )	Comments
Thread Plug Gages – Major Diameter Pitch Diameter	(0 to 12) in	$(50 + 0.6D) \mu\text{in}$	Sip measuring machine
Ring Gauges, Cylindrical & Tapered	(0.02 to 12) in	$(11 + 1.5L) \mu\text{in}$	Sip measuring machine
Indicators, Resolution –  0.00005 in 0.0001 in 0.001 in	(0 to 1) in (0 to 2) in (0 to 8) in	46 $\mu\text{in}$ 92 $\mu\text{in}$ 870 $\mu\text{in}$	Sip measuring machine
End Measuring Rods	(0 to 12) in	$(5 + 0.6L) \mu\text{in}$	Sip measuring machine
Pin Gages, Diameter	(0.02 to 2.00) in	75 $\mu\text{in}$	Sip measuring machine
Micrometers <sup>3</sup> – Length and Flatness, Resolution: 100 $\mu\text{in}$ 50 $\mu\text{in}$	(0.10 to 12) in	$(54 + 3L) \mu\text{in}$ $(28 + 2L) \mu\text{in}$	Mitutoyo gage blocks, optical flat, monochromic light source
Pitch Diameter, External Threads	(0.10 to 4.00) in	$(26 + 10L) \mu\text{in}$	Sip measuring machine

## II. Electrical – DC/Low Frequency

Parameter/Equipment	Range	Best Uncertainty <sup>2, 4, 5</sup> ( $\pm$ )	Comments
DC Voltage – Generate <sup>3</sup>	(0 to 220) mV 220 mV to 2.2 V (2.2 to 11) V (11 to 22) V (22 to 220) V (220 to 1100) V	11 $\mu\text{V/V} + 0.6 \mu\text{V}$ 10 $\mu\text{V/V} + 1.0 \mu\text{V}$ 11 $\mu\text{V/V} + 3.5 \mu\text{V}$ 10 $\mu\text{V/V} + 6.5 \mu\text{V}$ 11 $\mu\text{V/V} + 80 \mu\text{V}$ 13 $\mu\text{V/V} + 500 \mu\text{V}$	Fluke 5700A w/option 03

Parameter/Equipment	Range	Best Uncertainty <sup>2, 4, 5</sup> ( $\pm$ )	Comments
DC Voltage – Measure <sup>3</sup>	(0 to 100) mV 100 mV to 1V (1 to 10) V (10 to 100) V (100 to 1000) V	13 $\mu$ V/V + 3.0 $\mu$ V 17 $\mu$ V/V + 0.3 $\mu$ V 13 $\mu$ V/V + 0.5 $\mu$ V 15 $\mu$ V/V + 30 $\mu$ V 27 $\mu$ V/V + 100 $\mu$ V	HP 3458A
DC Current – Generate <sup>3</sup>	(0 to 220) $\mu$ A 220 $\mu$ A to 22 mA (22 to 220) mA 220 mA to 2.2 A (2.2 to 11) A  (11 to 20.5) A	50 $\mu$ A/A + 8 nA 50 $\mu$ A/A + 8 nA 50 $\mu$ A/A + 80 nA 60 $\mu$ A/A + 0.8 $\mu$ A 80 $\mu$ A/A + 25 $\mu$ A  0.12 %	Fluke 5700A w/option 03  Fluke 5500A
DC Current – Measure <sup>3</sup>	Up to 100 nA 100 nA to 1 $\mu$ A (1 to 10) $\mu$ A (10 to 100) $\mu$ A 100 $\mu$ A to 10 mA (10 to 100) mA 100 mA to 1 A	35 $\mu$ A/A + 400 $\mu$ A 25 $\mu$ A/A + 40 $\mu$ A 25 $\mu$ A/A + 10 $\mu$ A 25 $\mu$ A/A + 5 $\mu$ A 25 $\mu$ A/A + 5 $\mu$ A 40 $\mu$ A/A + 5 $\mu$ A 0.012 % + 10 $\mu$ A	HP 3458A
High DC Current <sup>3</sup>	(1 to 500) A	1.3 %	Fluke 5500A with current coil Hall Effect
Resistance – Generate <sup>3</sup>	Up to 11 $\Omega$ (11 to 33) $\Omega$ (33 to 110) $\Omega$ (110 to 330) $\Omega$ (0.33 to 1.1) k $\Omega$ (1.1 to 3.3) k $\Omega$ (3.3 to 11) k $\Omega$ (11 to 33) k $\Omega$ (33 to 110) k $\Omega$ (110 to 330) k $\Omega$ 0.33 k $\Omega$ to 1.1 M $\Omega$ (1.1 to 3.3) M $\Omega$ (3.3 to 11) M $\Omega$ (11 to 33) M $\Omega$ (33 to 110) M $\Omega$ (110 to 330) M $\Omega$ 330 M $\Omega$ to 1.1 G $\Omega$	0.12 % + 0.008 $\Omega$ 0.17 % 0.018 % 0.024 % 0.009 % + 0.06 $\Omega$ 0.024 % 0.009 % + 0.6 $\Omega$ 0.012 % 0.011 % + 6 $\Omega$ 0.013 % 0.015 % + 55 $\Omega$ 0.019 % 0.016 % 0.041 % 0.058 % 0.37 % 1.8 %	Fluke 5500A

Parameter/Equipment	Range	Best Uncertainty <sup>2,4,5</sup> ( $\pm$ )	Comments
Resistance (cont.) – Generate, Fixed Points <sup>3</sup>	1 $\Omega$ 10 $\Omega$ 100 $\Omega$ 1 k $\Omega$ 10 k $\Omega$ 100 k $\Omega$ 1 M $\Omega$ 10 M $\Omega$ 100 M $\Omega$	0.013 % 39 parts in 10 <sup>6</sup> 24 parts in 10 <sup>6</sup> 18 parts in 10 <sup>6</sup> 17 parts in 10 <sup>6</sup> 19 parts in 10 <sup>6</sup> 27 parts in 10 <sup>6</sup> 54 parts in 10 <sup>6</sup> 0.016 %	Fluke 5700A w/option 03
Resistance – Measure <sup>3</sup>	(0 to 10) $\Omega$ (10 to 100) $\Omega$ 100 $\Omega$ to 100 k $\Omega$ 100 k $\Omega$ to 1 M $\Omega$ (1 to 10) M $\Omega$ (10 to 100) M $\Omega$ 100 M $\Omega$ to 1 G $\Omega$	19 parts in 10 <sup>6</sup> + 0.06 m $\Omega$ 15 parts in 10 <sup>6</sup> + 0.6 m $\Omega$ 13 parts in 10 <sup>6</sup> + 0.6 m $\Omega$ 18 parts in 10 <sup>6</sup> + 2.4 $\Omega$ 59 parts in 10 <sup>6</sup> + 120 $\Omega$ 0.058 % + 1.2 k $\Omega$ 1.8 % + 10 k $\Omega$	HP 3458A
Capacitance – Generate <sup>3</sup>	(0.33 to 10.999) nF (11 to 109.99) nF (110 to 329.99) nF (0.33 to 1.0999) $\mu$ F (1.1 to 3.2999) $\mu$ F (3.3 to 10.999) $\mu$ F (11 to 32.999) $\mu$ F (33 to 109.99) $\mu$ F (110 to 329.99) $\mu$ F (330 to 1.1) mF	0.5 % + 0.01 nF 0.25 % + 0.1 nF 0.25 % + 0.3 nF 0.25 % + 1 nF 0.35 % + 3 nF 0.35 % + 10 nF 0.40 % + 30 nF 0.50 % + 100 nF 0.70 % + 300 nF 1 % + 300 nF	Fluke 5500A
Inductance – Generate <sup>3</sup>  Fixed Points	1.0 mH 100 mH 1 H	0.059 % rdg 0.062 % rdg 0.082 % rdg	Genrad 1482
Electrical Calibration of Thermocouple Indicators <sup>3</sup> –  Type E	-250 $^{\circ}$ C to -100 $^{\circ}$ C -100 $^{\circ}$ C to 650 $^{\circ}$ C 650 $^{\circ}$ C to 1000 $^{\circ}$ C	0.5 $^{\circ}$ C 0.16 $^{\circ}$ C 0.21 $^{\circ}$ C	Fluke 5500A

Parameter/Equipment	Range	Best Uncertainty <sup>2</sup> (±)	Comments
Electrical Calibration of Thermocouple Indicators (cont.) <sup>3</sup> –			
Type J	-210 °C to -100 °C -100 °C to 760 °C 760 °C to 1200 °C	0.27 °C 0.17 °C 0.23 °C	Fluke 5500A
Type K	-200 °C to -100 °C -100 °C to 120 °C 120 °C to 1000 °C 1000 °C to 1372 °C	0.33 °C 0.18 °C 0.26 °C 0.04 °C	
Type S	0 °C to 250 °C 250 °C to 1400 °C 1400 °C to 1767 °C	0.47 °C 0.37 °C 0.46 °C	
Type T	-250 °C to -150 °C -150 °C to 0 °C 0 °C to 400 °C	0.63 °C 0.24 °C 0.16 °C	
Electrical Calibration of RTD Indicating Systems <sup>3</sup> –			
Pt 385, 100 Ω	-200 °C to 0 °C 0 °C to 100 °C 100 °C to 400 °C 400 °C to 630 °C 630 °C to 800 °C	0.05 °C 0.07 °C 0.1 °C 0.12 °C 0.23 °C	Fluke 5500A
Pt 3926, 100 Ω	-200 °C to 0 °C 0 °C to 100 °C 100 °C to 400 °C 400 °C to 630 °C	0.05 °C 0.07 °C 0.1 °C 0.12 °C	
Pt 3916, 100 Ω	-200 °C to -190 °C -190 °C to 0 °C 0 °C to 300 °C 300 °C to 600 °C 600 °C to 630 °C	0.25 °C 0.05 °C 0.08 °C 0.1 °C 0.23 °C	
Pt 385, 200 Ω	-200 °C to 100 °C 100 °C to 260 °C 260 °C to 600 °C 600 °C to 630 °C	0.04 °C 0.05 °C 0.14 °C 0.16 °C	

Parameter/Equipment	Range	Best Uncertainty <sup>2</sup> (±)	Comments
Electrical Calibration of RTD Indicating Systems (cont.) <sup>3</sup> –			
Pt 385, 500 Ω	-200 °C to 100 °C 100 °C to 260 °C 260 °C to 600 °C 600 °C to 630 °C	0.05 °C 0.06 °C 0.09 °C 0.11 °C	Fluke 5500A
Pt 385, 1 kΩ	-200 °C to 100 °C 100 °C to 260 °C 260 °C to 600 °C 600 °C to 630 °C	0.03 °C 0.05 °C 0.07 °C 0.23 °C	
PtNi 385, 100 Ω	-80 °C to 100 °C 100 °C to 260 °C	0.08 °C 0.14 °C	
Cu 427, 10 Ω	-100 °C to 260 °C	0.3 °C	

Parameter/Range	Frequency	Best Uncertainty <sup>2,4</sup> (±)	Comments
AC Voltage – Generate <sup>3</sup>			
(0 to 220) mV	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz 100 kHz to 1 MHz	0.055 % + 13 μV 0.021 % + 8 μV 0.011 % + 8 μV 0.037 % + 8 μV 0.085 % + 25 μV 0.34 % + 80 μV	Fluke 5700A w/option 03
220 mV to 2.2 V	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz 100 kHz to 1 MHz	0.05 % + 80 μV 0.016 % + 25 μV 75 μV/V + 6 μV 0.012 % + 16 μV 0.025 % + 70 μV 0.22 % + 850 μV	
(2.2 to 22) V	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz 100 kHz to 1 MHz	0.05 % + 0.8 mV 0.016 % + 0.25 mV 75 μV/V + 0.06 mV 0.012 % + 0.16 mV 0.025 % + 0.35 mV 0.34 % + 8.5 mV	

Parameter/Range	Frequency	Best Uncertainty <sup>2, 4, 5</sup> ( $\pm$ )	Comments
AC Voltage (cont.) – Generate <sup>3</sup>			
(22 to 220) V	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz 100 kHz to 1 MHz	0.05 % + 8 mV 0.016 % + 2.5 mV 80 $\mu$ V/V + 0.8 mV 0.022 % + 3.5 mV 0.05 % + 8 mV 1.6 % + 190 mV	Fluke 5700A w/option 03
(220 to 750) V	(30 to 50) kHz (50 to 100) kHz	0.06 % + 11 mV 0.23 % 45 mV	
(220 to 1100) V	(15 to 50) Hz 50 Hz to 1 kHz	0.04 % + 16 mV 90 $\mu$ V/V + 4 mV	
AC Voltage – Measure <sup>3</sup>			
Up to 10 mV	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz	0.03 % + 3 $\mu$ V 0.02 % + 2 $\mu$ V 0.03 % + 2 $\mu$ V 0.12 % + 2 $\mu$ V 0.58 % + 2 $\mu$ V 4.6 % + 2 $\mu$ V	HP 3458A, synchronous sub- sampled mode
10 mV to 10 V	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz	80 $\mu$ V/V + 0.4 mV 80 $\mu$ V/V + 0.2 mV 0.02 % + 0.2 mV 0.03 % + 0.2 mV 0.09 % + 0.2 mV 0.35 % + 1 mV 1.2 % + 1 mV 1.7 % + 1 mV	
(10 to 100) V	(1 to 40) Hz 40 Hz to 1 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz	0.02 % + 4 mV 0.02 % + 2 mV 0.04 % + 2 mV 0.14 % + 2 mV 0.46 % + 10 mV 1.7 % + 10 mV	
(100 to 1000) V	(1 to 40) Hz 40 Hz to 1 kHz	0.05 % + 40 mV 0.05 % + 20 mV	
(1 to 20) kV	(1 to 20) kHz (20 to 50) kHz (50 to 100) kHz	0.07 % + 20 mV 0.14 % + 20 mV 0.35 % + 20 mV	

Parameter/Range	Frequency	Best Uncertainty <sup>2, 4, 5</sup> (±)	Comments
AC Current– Generate <sup>3</sup>  (1 to 220) µA 220 µA to 22 mA (22 to 220) mA 220 mA to 2.2 A	40 Hz to 1 kHz	0.09 % 0.024 % 0.026 % 0.093 %	Fluke 5700A w/option 03
AC Current <sup>3</sup> – Measure  Up to 100 µA  100 µA to 100 mA  100 mA to 1 A  (1 to 500) A	(10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz  (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz  (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz  (50 to 70) Hz	0.46 % + 0.03 µA 0.18 % + 0.03 µA 0.078 % + 0.03 µA  0.46 % + 20 µA 0.17 % + 20 µA 0.073 % + 20 µA 0.042 % + 20 µA  0.46 % + 200 µA 0.19 % + 200 µA 0.1 % + 200 µA 0.12 % + 200 µA  0.32 %	HP 3458A          Fluke 5500 with Fluke coil

III. Electrical – RF/Microwave

Parameter/Range	Frequency	Best Uncertainty <sup>2</sup> (±)	Comments
RF Tuned Power Relative – Measure <sup>3</sup>  (-0.0 to -3) dB (-3 to -10) dB (-10 to -40) dB (-40 to -50) dB (-50 to -80) dB (-80 to -90) dB (-90 to -110) dB (-110 to -127) dB	2.5 MHz (18 to 26.5)	0.02 dB 0.02 dB 0.08 dB 0.14 dB 0.2 dB 0.26 dB 0.3 dB 0.4 dB	HP 8902A, HP11722A,

Parameter/Range	Frequency	Best Uncertainty <sup>2</sup> (±)	Comments
RF Power <sup>3</sup> – Absolute RF, Measure (-30 to -20 dBm (10 to 20) dBm	(50 MHz to 50 GHz) (50 MHz to 50 GHz)	0.06 dBm 1.2 dB	HP 438A, 8484A, 8482A 8485A, 8487A
Amplitude Modulation – Measure <sup>3</sup>  Rate: 150 kHz to 10 MHz Depth: (5 to 99) %  Rate: 10 MHz to 1.3 GHz Depth: (5 to 99) %	50 Hz to 10 kHz 20 Hz to 10 kHz  50 Hz to 10 kHz 20 Hz to 10 kHz	2 % + 1 digit of rdg 3 % + 1 digit of rdg  1 % + 1 digit of rdg 3 % + 1 digit of rdg	HP 8902A
Frequency Modulation – Measure <sup>3</sup>  Rate: 250 kHz to 10 MHz Dev: ≤ 40 kHz  Rate: 10 MHz to 1.3 GHz Dev: ≤ 400 kHz	50 Hz to 10 kHz  50 Hz to 100 kHz 20 Hz to 200 kHz	2 % + 1 digit of rdg  1 % + 1 digit of rdg 5 % + 1 digit of rdg	HP 8902A

#### IV. Mechanical

Parameter/Equipment	Range	Best Uncertainty <sup>2</sup> (±)	Comments
Torque <sup>3</sup> – Measure	(0 to 400) in·oz (0 to 50) in·lb (0 to 150) in·lb (0 to 400) in·lb (0 to 600) ft·lb	0.83 % of rdg 0.92 % of rdg 0.7 % of rdg 0.66 % of rdg 0.68 % of rdg	CDI torque system
Scales and Balances <sup>3</sup>	1mg to 30 kg  (10 to 1000) lb	1.0 LSVD  300 mg	Verification with class 1  Class F weights

Parameter/Equipment	Range	Best Uncertainty <sup>2</sup> (±)	Comments
Mass –	(1 to 500) mg	2 µg	Class 1 with balance
Fixed Points	1 g 2 g 5 g 10 g 20 g 50 g 100 g 200 g  10 kg	3 µg 4 µg 5 µg 8 µg 10 µg 23 µg 36 µg 44 µg  3 mg	Class 1 with balance          Class 4 with balance
Pressure – Measure <sup>3</sup>	(-15 to 100) psi Up to 1000 psi Up to 10 000 psi	0.1 % of rdg 0.09 % of rdg 0.15 % of rdg	Heise PTE-1 Fluke 700P08 Heise HQS-2-B-I-G

#### V. Thermodynamics

Parameter/Equipment	Range	Best Uncertainty <sup>2</sup> (±)	Comments
Humidity – Measuring Equipment <sup>3</sup>	11 % RH 33 % RH 75 % RH 97 % RH	1.6 % RH of rdg 1.7 % RH of rdg 1.5 % RH of rdg 2 % RH of rdg	Saturated salt solutions
Temperature – Measuring Equipment	(-20 to 600) °C	0.53 °C	SPRT and Hart 9105 calibrator

#### VI. Time & Frequency

Parameter/Equipment	Range	Best Uncertainty <sup>2</sup> (±)	Comments
Frequency – Measuring Equipment	10 MHz	5 x 10 <sup>-12</sup> Hz	HP 58503A, GPS

---

<sup>1</sup> This laboratory offers commercial calibration service and field calibration service.

<sup>2</sup> “Best Uncertainty” is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards of nearly ideal measuring equipment. Best uncertainties represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The best uncertainty of a specific calibration performed by the laboratory may be greater than the best uncertainty due to the behavior of the customer’s device and to influences from the circumstances of the specific calibration.

<sup>3</sup> Field calibration service is available for this calibration and this laboratory meets A2LA R104 – *General Requirements: Accreditation of Field Testing and Field Calibration Laboratories* for these calibrations. Please note the uncertainties achievable on a customer's site can normally be expected to be larger than the Best Measurement Capabilities (BMC) that the accredited laboratory has been assigned as Best Uncertainty on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the calibration uncertainty being larger than the BMC.

<sup>4</sup> The measurands stated are generated with the Fluke 5500A, Fluke 5700A series of instruments. This capability is suitable for the calibration of the devices intended to measure the stated measurand in the ranges indicated. Best measurement uncertainties are expressed as either a specific value that covers the full range or as a fraction of the reading plus a fixed floor specification.

<sup>5</sup> The measurands stated are measured with the HP 3458A series of instruments. This capability is suitable for the calibration of the devices intended to generate the measurand in the ranges indicated. Best measurement uncertainties are expressed as either a specific value that covers the full range or as a combination of the fraction of the reading/output plus a range specification.

<sup>6</sup> In the statement of best uncertainty,  $L$  is the numerical value of the nominal length of the device measured in inches.  $D$  is the numerical value of the nominal diameter of the device measured in inches. Pitch diameter is measured by the three-wire method.



World Class Accreditation

The American Association for Laboratory Accreditation

# Accredited Laboratory

A2LA has accredited

## MICRO PRECISION CALIBRATION, INC.

*Tampa, FL*

for technical competence in the field of

### Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 *General Requirements for the Competence of Testing and Calibration Laboratories*. This laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and any additional program requirements in the field of calibration. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (*refer to joint ISO-ILAC-IAF Communiqué dated 8 January 2009*).

Presented this 2nd day of June 2009.



A handwritten signature in black ink, appearing to read "Peter Meyer".

President & CEO  
For the Accreditation Council  
Certificate Number 935.09  
Valid to July 31, 2011

*For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.*