



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005,
ANSI/NCSL Z540-1-1994 & ANSI/NCSL Z540.3-2006

AGILENT TECHNOLOGIES LOCAL CALIBRATION CENTER
 8000 W Sunrise Blvd
 Plantation, FL 33322
 Ron Quinn Phone: 321 235 2149

CALIBRATION

Valid To: July 31, 2013

Certificate Number: 2600.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations¹

I. Electrical – DC/Low Frequency

Parameter/Equipment	Range	CMC ^{2,4,5} (±)	Comments
DC Voltage – Generate ³	(0 to 0.22) V (0.22 to 2.2) V (2.2 to 11) V (11 to 22) V (22 to 220) V (220 to 1100) V	3.8 μV/V + 2.5 μV 4.9 μV/V + 1.9 μV 3.3 μV/V + 4.7 μV 2.4 μV/V + 48 μV 5.4 μV/V + 40 μV 6.9 μV/V + 520 μV	Fluke 5720A/5725A
DC Voltage – Measure ³	(0 to 0.1) V (0.1 to 1) V (1 to 10) V (10 to 100) V (100 to 1000) V	5.6 μV/V + 1.5 μV 5.2 μV/V + 1.2 μV 4.7 μV/V + 2.5 μV 6.6 μV/V + 45 μV 19 μV/V + 160 μV	HP 3458A/100 PLC option 002

Parameter/Equipment	Range	CMC ^{2, 4, 5, 11} (\pm)	Comments
DC Current – Generate ³	(0 to 220) μ A 220 μ A to 2.2 mA (2.2 to 22) mA (22 to 100) mA (100 to 220) mA 220 mA to 1 A (1 to 2.2) A (2.2 to 11) A	36 μ A/A + 0.009 μ A 33 μ A/A + 0.008 μ A 34 μ A/A + 0.040 μ A 42 μ A/A + 0.68 μ A 51 μ A/A + 0.00 μ A 76 μ A/A + 12 μ A 0.015 % - 66 μ A 0.030 % + 400 μ A	Fluke 5720A
DC Current – Measure ³	(10 to 100) μ A 100 μ A to 1 mA (1 to 10) mA (10 to 100) mA 100 mA to 1.1 A	20 μ A/A + 0 μ A 20 μ A/A + 0.1 μ A 20 μ A/A + 0.06 μ A 35 μ A/A + 0.60 μ A 0.011 % + 11 μ A	HP 3458A
Shunt 15 A 0.1 Ω , 25 W	Dissipated power: 0.3 Amps to full power	0.014 %	Guildline 9230-15
Shunt 100A 0.01 Ω , 100 W	\leq 25W	0.014 %	Guildline 9230-100
Shunt 300A 0.01 Ω , 90 W	\leq 25W	0.012 %	Guildline 9230-300
Resistance – Generate, Fixed Points ³	0 Ω 1 Ω 1.9 Ω 10 Ω 19 Ω 100 Ω 190 Ω 1 k Ω 1.9 k Ω 10 k Ω 19 k Ω 100 k Ω 190 k Ω 1 M Ω 1.9 M Ω 10 M Ω 19 M Ω 100 M Ω	0.25 m Ω 0.27 m Ω 0.31 m Ω 0.34 m Ω 2.5 m Ω 2.7 m Ω 3.3 m Ω 9.3 m Ω 31 m Ω 93 m Ω 0.19 Ω 1.2 Ω 2.2 Ω 20 Ω 42 Ω 400 Ω 1.5 k Ω 12 k Ω	Fluke 5720A

Parameter/Equipment	Range	CMC ^{2,4,5} (±)	Comments
Resistance – Measure ³	(0 to 10) Ω (10 to 100) Ω 100 Ω to 1 kΩ (1 to 10) kΩ (10 to 100) kΩ 100 kΩ to 1 MΩ (1 to 10) MΩ (10 to 100) MΩ 100 MΩ to 1 GΩ	18 μΩ/Ω + 74 μΩ 13 μΩ/Ω + 740 μΩ 11 μΩ/Ω + 7.6 μΩ 11 μΩ/Ω + 74 μΩ 11 μΩ/Ω + 740 μΩ 15 μΩ/Ω + 2.4 μΩ 53 μΩ/Ω + 110 μΩ 0.05 % + 4.3 kΩ 0.5 % + 290 kΩ	HP 3458A

Parameter/Range	Frequency	CMC ^{2,4,5} (±)	Comments
AC Current – Generate ³			
(0 to 220) μA	(10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.016 % + 0.063 μA 89 μA/A + 0.062 μA 60 μA/A + 0.062 μA 0.018 % + 0.062 μA 0.10 % + 0.09 μA	Fluke 5720A
220 μA to 2.2 mA	(10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.027 % + 0.056 μA 0.020 % + 0.051 μA 0.016 % + 0.052 μA 0.023 % + 0.12 μA 0.11 % + 0.67 μA	
(2.2 to 22) mA	(10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.027 % + 0.56 μA 0.021 % + 0.51 μA 0.016 % + 0.52 μA 0.023 % + 0.71 μA 0.11 % + 5.1 μA	
(22 to 220) mA	(10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.028 % + 3.9 μA 0.021 % + 3.0 μA 0.017 % + 2.2 μA 0.024 % + 3.1 μA 0.11 % + 0.01 mA	
220 mA to 2.2 A	20 Hz to 1 kHz 220 mA to 1.0 A >1 A to 2.2 A (1 to 5) kHz 220 mA to 1.0 A >1 A to 2.2 A (5 to 10) kHz	0.030 % + 32 μA 0.035 % + 26 μA 0.044 % + 83 μA 0.055 % + 72 μA 0.67 % + 0.072 mA	

Peter Abney

Parameter/Range	Frequency	CMC ^{2,4,5} (±)	Comments
AC Current – Generate ³ (cont) (2.2 to 11) A	40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.044 % + 130 µA 0.088 % + 290 µA 0.31 % + 0.64 mA	Fluke 5720A
AC Current – Measure ³ (0 to 100) µA (0.1 to 1) mA (1 to 10) mA (10 to 100) mA 100 mA to 1.05 A	(10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz (10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz (10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz (10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz	0.4 % + 31 nA 0.15 % + 31 nA 0.06 % + 31 nA 0.4 % + 310 nA 0.15 % + 210 nA 0.06 % + 210 nA 0.4 % + 3.1 µA 0.15 % + 2.1 µA 0.06 % + 2.1 µA 0.4 % + 31 µA 0.15 % + 21 µA 0.06 % + 21 µA 0.4 % + 220 µA 0.16 % + 220 µA 0.08 % + 220 µA 0.1 % + 220 µA	HP 3458A

Parameter/Range	Frequency	CMC ^{2,4,5} (±)	Comments
AC Voltage – Generate ³			
(0 to 2.2) mV	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.027 % + 4.5 μV 0.013 % + 4.5 μV 0.011 % + 4.5 μV 0.027 % + 4.5 μV 0.054 % + 5.3 μV 0.12 % + 10 μV 0.17 % + 24 μV 0.30 % + 24 μV	
(22 to 220) mV	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.025 % + 12 μV 95 μV/V + 7.3 μV 86 μV/V + 7.3 μV 0.018 % + 9.4 μV 0.049 % + 16 μV 0.088 % + 20 μV 0.14 % + 0.033 mV 0.27 % + 0.048 mV	
220 mV to 2.2 V	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.025 % + 39 μV 99 μV/V + 15 μV 63 μV/V + 5.9 μV 86 μV/V + 8.3 μV 0.011 % + 30 μV 0.041 % + 78 μV 0.99 % + 200 μV 0.16 % + 0.32 mV	
(2.2 to 22) V	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.026 % + 420 μV 0.011 % + 150 μV 67 μV/V + 44 μV 91 μV/V + 90 μV 0.011 % + 210 μV 0.029 % + 650 μV 0.11 % + 2.0 mV 0.16 % + 3.3 mV	
(22 to 220) V	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (>22 to 100) Vrms (>100 to 220) Vrms (50 to 100) kHz (>22 to 100) Vrms (>100 to 220) Vrms (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.024 % + 3.9 mV 0.010 % + 1.5 mV 69 μV/V + 430 μV 98 μV/V + 910 μV 0.017 % + 560 μV 0.016 % + 2.4 mV 0.021 % + 1.9 mV 0.087 % + 16 mV 0.43 % + 39 mV 0.79 % + 79 mV	

Peter Abney

Parameter/Range	Frequency	CMC ^{2,4,5} (±)	Comments
AC Voltage – Generate ³ (cont)			
(0 to 250) V max output, 1100 V range	(15 to 50) Hz 50 Hz to 1 kHz (1 to 20) kHz (20 to 30) kHz	0.031 % + 17 mV 91 µV/V + 2.9 mV 91 µV/V + 2.9 mV 0.051 % + 9.6 mV	Fluke 5720A/ 5725A
(0 to 750) V	(30 to 50) kHz (50 to 100) kHz	0.052 % + 8.6 mV 0.19 % + 37 mV	
AC Voltage – Measure ³			
(0 to 10) mV	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 4) MHz (4 to 8) MHz	0.03 % + 3.1 µV 0.02 % + 1.2 µV 0.03 % + 1.7 µV 0.1 % + 1.6 µV 0.5 % + 1.3 µV 4 % + 2.1 µV 1.2 % + 6.6 µV 7 % + 7.5 µV 20 % + 8.2 µV	HP 3458A
(10 to 100) mV	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz (2 to 4) MHz (4 to 8) MHz (8 to 10) MHz	70 µV/V + 4.1 µV 70 µV/V + 2.1 µV 0.014 % + 2.3 µV 0.03 % + 2.6 µV 0.08 % + 2.3 µV 0.3 % + 15 µV 1 % + 28 µV 1.5 % + 20 µV 4 % + 74 µV 4 % + 83 µV 15 % + 110 µV	
100 mV to 1 V	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz (2 to 4) MHz (4 to 8) MHz (8 to 10) MHz	70 µV/V + 41 µV 70 µV/V + 21 µV 0.014 % + 22 µV 0.03 % + 22 µV 0.08 % + 22 µV 0.3 % + 120 µV 1 % + 300 µV 1.5 % + 210 µV 4 % + 730 µV 4 % + 830 µV 15 % + 1 mV	

Peter Abney

Parameter/Range	Frequency	CMC ^{2,4,5} (±)	Comments
AC Voltage – Measure ³			
(1 to 10) V	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz (2 to 4) MHz (4 to 8) MHz (8 to 10) MHz	70 µV/V + 420 µV 70 µV/V + 220 µV 0.014 % + 240 µV 0.03 % + 250 µV 0.08 % + 220 µV 0.3 % + 1.1 mV 1 % + 1.1 mV 1.5 % + 1.1 mV 4 % + 7.1 mV 4 % + 8.1 mV 15 % + 11 mV	HP 3458A
(10 to 100) V	(1 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz	0.02 % + 4.1 mV 0.02 % + 2.6 mV 0.035 % + 2.4 mV 0.12 % + 2.1 mV 0.4 % + 11 mV 1.5 % + 40 mV	
(100 to 750) V	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz	0.04 % + 31 mV 0.04 % + 16 mV 0.06 % + 16 mV 0.12 % + 16 mV 0.3 % + 15 mV	
AC Voltage Flatness – Generate ³			
0.3 mV to 3.5 V	(10 to 30) Hz 30 Hz to 120 kHz	0.27 % 0.14 %	Fluke 5720A or 5720A-03 (referenced to 1 kHz)
(0.3 to 1.1) mV	120 kHz to 2 MHz (2 to 10) MHz (10 to 20) MHz (20 to 30) MHz	0.46 % 0.62 % 0.80 % 2.4 %	
(0.3 to 3) mV	120 kHz to 2 MHz (2 to 10) MHz (10 to 20) MHz (20 to 30) MHz	0.22 % 0.37 % 0.55 % 1.4 %	
3 mV to 3.5 V	120 kHz to 2 MHz (2 to 10) MHz (10 to 20) MHz (20 to 30) MHz	0.12 % 0.21 % 0.38 % 0.86 %	

Parameter/Range	Frequency	CMC ^{2, 11} (±)	Comments
AC Voltage Flatness – Measure ³ Up to 3 V	10 Hz 100 Hz (10, 30) kHz 100 kHz 300 kHz 1 MHz 3 MHz 8 MHz 10 MHz 20 MHz 30 MHz 50 MHz 70 MHz 80 MHz 100 MHz	0.02 % + 6.9 μV/V 80 μV + 5.5 μV/V 80 μV + 3.2 μV/V 0.01 % + 8.0 μV/V 0.01 % + 5.2 μV/V 0.01 % + 6.5 μV/V 0.13 % + 59 μV/V 0.13 % + 110 μV/V 0.13 % + 91 μV/V 0.25 % + 210 μV/V 0.25 % + 240 μV/V 0.61 % + 340 μV/V 0.9 % + 240 μV/V 1.1 % + 790 μV/V 1.3 % + 940 μV/V	By comparison to 11049A, 11050A, 11051A thermal voltage converters
Resistance – Generate ³ 0.1 Ω (1, 10) Ω 100 Ω (1, 10, 100) kΩ	DC to 1 MHz, direct measurement	1 % of value 0.1 % of value 0.03 % of value 0.03 % of value	16074A DUT box calibration R-L standard, BNC 3 terminal
Capacitance – Generate ³ Direct Measure (1, 10) pF (100, 1000) pF (0.01, 0.1, 1) μF	1 kHz	0.01 % 0.01 % 0.01 %	16380A/16380C standard air capacitor set, BNC 4 terminal pair

Parameter/Range	Frequency	CMC ^{2, 11} (±)	Comments
Capacitance – Generate ³ (cont)			
Algorithmic Derivation 1 pF	1 MHz 2 MHz 3 MHz 4 MHz 5 MHz 10 MHz 13 MHz	0.05 % 0.06 % 0.10 % 0.2 % 0.3 % 1.0 % 1.5 %	16380A/16380C standard air capacitor set, BNC 4 terminal pair
(10, 100) pF	(1, 2) MHz 3 MHz 4 MHz 5 MHz 10 MHz 13 MHz	0.025 % 0.03 % 0.04 % 0.06 % 0.15 % 0.20 %	
1000 pF	1 MHz 2 MHz 3 MHz 4 MHz 5 MHz 10 MHz 13 MHz	0.05 % 0.06 % 0.10 % 0.15 % 0.20 % 0.50 % 0.70 %	
Substitution Method (0.01, 0.1, 1) μF	120 Hz to 10 kHz	0.025 % 0.04 %	
(0.01, 0.1) μF	100 kHz	0.05 %	
1 μF		0.1 %	

II. Electrical – RF/Microwave

Parameter/Equipment/Range	CMC ^{2, 10} (±)	Comments
<p>Amplitude Modulation – Measure^{3, 8}</p> <p>(0.15 to 10) MHz (5 % to < 10 %) AM (10 % to 99 %) AM</p> <p>(5 % to < 10 %) AM (10 % to 99 %) AM</p> <p>(10 to 1300) MHz (5 % to < 10 %) AM (10 % to 99 %) AM</p> <p>(5 % to < 10 %) AM (10 % to 99 %) AM</p> <p>1300 MHz to 26.5 GHz (5 % to < 10 %) AM (10 % to 99 %) AM</p> <p>10 MHz to 26.5 GHz (5 % to < 10 %) AM (10 % to 99 %) AM</p>	<p>0.025AM + 0.030 % 0.025AM + 0.14 %</p> <p>0.038AM + 0.030 % 0.038AM + 0.13 %</p> <p>0.012AM + 0.033 % 0.012AM + 0.17 %</p> <p>0.038AM + 0.030 % 0.037AM + 0.16 %</p> <p>0.019AM + 0.029 % 0.019AM + 0.14 %</p> <p>0.038AM + 0.030 % 0.038AM + 0.11 %</p>	<p>HP 8902 measuring receiver:</p> <p>rate: 50 Hz to 10 kHz, depth: 5 % to 99 %</p> <p>rate: 20 Hz to 10 kHz, depth: to 99 %</p> <p>rate: 50 Hz to 10 kHz, depth: 5 % to 99 %</p> <p>rate: 20 Hz to 10 kHz, depth: to 99 %</p> <p>rate: 50 Hz to 10 kHz, depth: 5 % to 99 %</p> <p>rate: 20 Hz to 10 kHz, depth: to 99 %</p>
<p>Frequency Modulation – Measure^{3, 9}</p> <p>(0.25 to 10) MHz (0 to < 4) kHz Peak FM (≥ 4 to < 40) kHz Peak FM</p> <p>(10 to 1300) MHz (0 to < 4) kHz Peak FM (≥ 4 to < 40) kHz Peak FM (≥ 40 to < 400) kHz Peak FM</p> <p>(10 to 1300) MHz (0 to < 4) kHz Peak FM (≥ 4 to < 40) kHz Peak FM (≥ 40 to < 400) kHz Peak FM</p>	<p>0.024FM + 2.6 Hz Pk 0.024FM + 10 Hz Pk</p> <p>0.012FM + 2.6 Hz Pk 0.012FM + 12 Hz Pk 0.012FM + 110 Hz Pk</p> <p>0.059FM + 2.8 Hz Pk 0.059FM + 13 Hz Pk 0.059FM + 120 Hz Pk</p>	<p>HP 8902 measuring receiver: 40.0 to 400.0: 1 digit = 100 Hz 4.00 to 39.99: 1 digit = 10 Hz 0 to 3.999: 1 digit = 1 Hz</p> <p>rate: 20 Hz to 10 kHz, ≤ 40 kHz peak</p> <p>rate: 50 Hz to 100 kHz, ≤ 400 kHz peak</p> <p>rate: 20 Hz to 200 kHz, ≤ 400 kHz peak</p>

Parameter/Equipment/Range	CMC ^{2, 10} (±)	Comments
<p>Frequency Modulation – Measure^{3, 9} (cont)</p> <p>10 MHz to 26.5 GHz (10 to 1300) MHz (0 to < 4) kHz Peak FM (≥ 4 to < 40) kHz Peak FM (≥ 40 to < 400) kHz Peak FM</p> <p>(> 1.3 to 6.2) GHz (0 to < 4) kHz Peak FM (≥ 4 to < 40) kHz Peak FM (≥ 40 to < 400) kHz Peak FM</p> <p>(> 6.2 to 12.4) GHz (0 to < 4) kHz Peak FM (≥ 4 to < 40) kHz Peak FM (≥ 40 to < 400) kHz Peak FM</p> <p>(> 12.4 to 18.6) GHz (0 to < 4) kHz Peak FM (≥ 4 to < 40) kHz Peak FM (≥ 40 to < 400) kHz Peak FM</p> <p>(> 18.6 to 26.5) GHz (0 to < 4) kHz Peak FM (≥ 4 to < 40) kHz Peak FM (≥ 40 to < 400) kHz Peak FM</p> <p>10 MHz to 26.5 GHz (10 to 1300) MHz (0 to < 4) kHz Peak FM (≥ 4 to < 40) kHz Peak FM (≥ 40 to < 400) kHz Peak FM</p> <p>(> 1.3 to 6.2) GHz (0 to < 4) kHz Peak FM (≥ 4 to < 40) kHz Peak FM (≥ 40 to < 400) kHz Peak FM</p> <p>(> 6.2 to 12.4) GHz (0 to < 4) kHz Peak FM (≥ 4 to < 40) kHz Peak FM (≥ 40 to < 400) kHz Peak FM</p> <p>(> 12.4 to 18.6) GHz (0 to < 4) kHz Peak FM (≥ 4 to < 40) kHz Peak FM (≥ 40 to < 400) kHz Peak FM</p>	<p>0.012FM + 2.4 Hz Pk 0.012FM + 12.0 Hz Pk 0.012FM + 100 Hz Pk</p> <p>0.0099FM + 10 Hz Pk 0.012FM + 12 Hz Pk 0.012FM + 100 Hz Pk</p> <p>0.0075FM + 24 Hz Pk 0.012FM + 12 Hz Pk 0.012FM + 85 Hz Pk</p> <p>0.0049FM + 52 Hz Pk 0.011FM + 36 Hz Pk 0.012FM + 110 Hz Pk</p> <p>0.0035FM + 80 Hz Pk 0.011FM + 46 Hz Pk 0.012FM + 100 Hz Pk</p> <p>0.059FM + 2.8 Hz Pk 0.059FM + 14 Hz Pk 0.059FM + 120 Hz Pk</p> <p>0.058FM + 5.2 Hz Pk 0.059FM + 14 Hz Pk 0.059FM + 120 Hz Pk</p> <p>0.055FM + 15 Hz Pk 0.059FM + 14 Hz Pk 0.059FM + 120 Hz Pk</p> <p>0.050FM + 37 Hz Pk 0.059FM + 15 Hz Pk 0.059FM + 120 Hz Pk</p>	<p>HP 8902 measuring receiver: 11793A DownConverter, Gen. Ext L.O.</p> <p>rate: 50 Hz to 100 kHz, ≤ 400 kHz peak</p> <p>rate: 20 Hz to 200 kHz, ≤ 400 kHz peak</p>

Parameter/Equipment/Range	CMC ^{2,10} (±)	Comments
Frequency Modulation – Measure ^{3,9} (cont) (> 18.6 to 26.5) GHz (0 to < 4) kHz Peak FM (≥ 4 to < 40) kHz Peak FM (≥ 40 to < 400) kHz Peak FM	 0.045FM + 62 Hz Pk 0.059FM + 16 Hz Pk 0.059FM + 120 Hz Pk	HP 8902 measuring receiver: 11793A DownConverter, Gen. Ext L.O. rate: 20 Hz to 200 kHz, ≤400 kHz peak

Parameter/Range	Frequency	CMC ² (±)	Comments
Digital Modulation – Measure ³ Carrier: 2 MHz to 2.65 GHz Error Vector Magnitude for Modulation Types: MSK, GMSK, BPSK, DQPSK, π/4DQPSK, 8PSK, 16QAM and 32QAM, QPSK Phase Error for Modulation Types: MSK, GMSK, BPSK, DQPSK, π/4DQPSK, 8PSK, 16QAM and 32QAM, QPSK Error Vector Magnitude for FSK Modulation	Mod Frequency Span: (1 to 100) kHz (0.1 to 1) MHz 1 MHz to 2.65 GHz (1 to 100) kHz (0.1 to 1) MHz 1 MHz to 2.65 GHz Mod Frequency: 3.2 kHz 1.152 kHz	 0.33 % rms 0.51 % rms 1.0 % rms 0.18° rms 0.34° rms 0.57° rms 0.54 % rms 1.5 % rms	 HP 89441A vector signal analyzer

Parameter/Range	Frequency	CMC ² (±)	Comments
RF Absolute Power – Measure ^{3,6}			
1 mW, Type-N(f), 50 Ω	50 MHz	0.015 dB (3.4 μW)	HP 432A w/ HP 478A-H76
(+20 to -30) dBm, 75 Ω	(100 to 600) kHz SWR ≤ 1.8:1	0.057 dB	HP 436A or HP 438A w/ HP 8483A, Type-N(m)
	600 kHz to 2 GHz SWR ≤ 1.18:1	0.059 dB	
(+20 to -30) dBm, 50 Ω	(100 to 300) kHz SWR ≤ 1.6:1	0.05 dB	HP 436A or HP 438A w/ HP 8482A, Type-N(m)
	300 kHz to 1 MHz SWR ≤ 1.2:1	0.049 dB	
	1 MHz to 2 GHz SWR ≤ 1.1:1	0.051 dB	
	(2 to 4.2) GHz SWR ≤ 1.3:1	0.050 dB	
(-20 to -70) dBm, 50 Ω	(10 to 30) MHz SWR ≤ 1.4:1	0.074 dB	HP 436A or HP 438A w/ HP 8481D, Type-N(m)
	30 MHz to 4 GHz SWR ≤ 1.15:1	0.074 dB	
	(4 to 10) GHz SWR ≤ 1.21:1	0.077 dB	
	(10 to 15) GHz SWR ≤ 1.3:1	0.10 dB	
	(15 to 18) GHz SWR ≤ 1.35:1	0.11 dB	

Parameter/Range	Frequency	CMC ² (±)	Comments
RF Absolute Power – Measure ^{3, 6} (cont) (+20 to -30) dBm, 50 Ω	(50 to 100) MHz SWR ≤ 1.15:1	0.060 dB	HP 436A or HP 438A w/ HP 8487A, 2.4 mm(m)
	(0.1 to 2) GHz SWR ≤ 1.1:1	0.075 dB	
	(2 to 12.4) GHz SWR ≤ 1.15:1	0.062 dB	
	(12.4 to 18) GHz SWR ≤ 1.2:1	0.065 dB	
	(18 to 26.5) GHz SWR ≤ 1.25:1	0.099 dB	
	(26.5 to 40) GHz SWR ≤ 1.4:1	0.097 dB	
	(40 to 50) GHz SWR ≤ 1.5:1	0.13 dB	
(-20 to -70) dBm, 50 Ω	(50 to 100) MHz SWR ≤ 1.19:1	0.054 dB	HP 436A or HP 438A w/ HP 8487D, 2.4 mm(m)
	(0.1 to 2) GHz SWR ≤ 1.15:1	0.054 dB	
	(2 to 12.4) GHz SWR ≤ 1.2:1	0.058 dB	
	(12.4 to 18) GHz SWR ≤ 1.29:1	0.068 dB	
	(18 to 34) GHz SWR ≤ 1.37:1	0.095 dB	
	(34 to 40) GHz SWR ≤ 1.61:1	0.11 dB	
	(40 to 50) GHz SWR ≤ 1.86:1	0.19 dB	
(+30 to -20) dBm, 50 Ω	100 kHz to 2.6 GHz SWR ≤ 1.15:1	0.071 dB	HP 8902A w/ HP 11722A, Type-N(m)
(+30 to -20) dBm, 50 Ω	(50 to 1300) MHz SWR ≤ 1.15:1	0.071 dB	HP 8902A w/ HP 11792A, APC 3.5 mm(m)
	(1.3 to 18) GHz SWR ≤ 1.25:1	0.081 dB	
	(18 to 26.5) GHz SWR ≤ 1.4:1	0.092 dB	

Parameter/Range	Frequency	CMC ² (±)	Comments		
Tuned RF Power ³ – Absolute – Measure ⁷	2.5 MHz to 26.5 GHz		HP 8902A with HP 11722A or with HP 11792A and HP 11793A		
(≤ +10 to ≥ -22) dBm		0.17 dB			
(< -22 to ≥ -42) dBm		0.18 dB			
(< -42 to ≥ -50) dBm		0.20 dB			
(< -50 to ≥ -60) dBm		0.21 dB			
(< -60 to ≥ -72) dBm		0.22 dB			
(< -72 to ≥ -80) dBm		0.23 dB			
(< -80 to ≥ -92) dBm		0.24 dB			
(< -92 to ≥ -102) dBm		0.27 dB			
(< -102 to ≥ -110) dBm		0.28 dB			
(< -110 to ≥ -120) dBm		0.31 dB			
(< -120 to ≥ -127) dBm		0.34 dB			
Relative – Measure		2.5 MHz to 26.5 GHz			HP 8902A with HP 11722A or with HP 11792A and HP 11793A
(≤ +10 to ≥ +2) dBm				0.081 dB	
(< + 2 to ≥ -12) dBm	0.071 dB				
(< -12 to ≥ -22) dBm	0.081 dB				
(< -22 to ≥ -31) dBm	0.088 dB				
(< -31 to ≥ -40) dBm	0.095 dB				
(< -40 to ≥ -50) dBm	0.12 dB				
(< -50 to ≥ -61) dBm	0.15 dB				
(< -61 to ≥ -71) dBm	0.16 dB				
(< -71 to ≥ -80) dBm	0.17 dB				
(< -80 to ≥ -90) dBm	0.19 dB				
(< -90 to ≥ -100) dBm	0.22 dB				
(< -100 to ≥ -110) dBm	0.23 dB				
(< -110 to ≥ -120) dBm	0.27 dB				
(< -120 to ≥ -127) dBm	0.30 dB				

Parameter/Range	Frequency	CMC ² (±)	Comments
RF Absolute Power – Generate ^{3,6}			
Into 50 Ω (10 to 3) V p-p	0.001 Hz to 100 kHz SWR 1.2:1	0.12 dB	HP 3325A/B with BNC(f)
2.99 V to 1 mV p-p	0.001 Hz to 100 kHz SWR 1.2:1	0.23 dB	
Into 50 Ω (10 to 3) V p-p	100 kHz to 20 MHz SWR ≤ 1.2:1	0.47 dB	HP 3325A/B with BNC(f)
2.999 V to 1 mV p-p	100 kHz to 10 MHz SWR ≤ 1.2:1	0.70 dB	
(2.999 to 0.1) V p-p	(10 to 20) MHz SWR ≤ 1.2:1	0.70 dB	
(99.99 to 1) mV p-p	(10 to 20) MHz SWR ≤ 1.2:1	1.0 dB	
Full Amplitude, 50 Ω 13.01 dBm	1 kHz to 25 MHz 200 Hz to 80 MHz	0.14 dB 0.23 dB	HP 3335A/B with BNC(f)
Full Amplitude, 75 Ω 11.25 dBm	1 kHz to 25 MHz 200 Hz to 25 MHz	0.14 dB 0.23 dB	HP 3335A/B with BNC(f)
In 2 dB steps, 50 Ω (0 to -18) dBm (-20 to -58) dBm (-60 to -98) dBm	200 Hz to 80 MHz 200 Hz to 80 MHz 200 Hz to 80 MHz	0.28 dB 0.34 dB 0.47 dB	HP 3335A/B with BNC(f)
In 2 dB steps, 75 Ω (0 to -18) dBm	200 Hz to 25 MHz (25 to 80) MHz	0.28 dB 0.41 dB	HP 3335A/B with BNC(f)
(-20 to -58) dBm	200 Hz to 25 MHz (25 to 80) MHz	0.34 dB 0.52 dB	
(-60 to -98) dBm	200 Hz to 25 MHz (25 to 80) MHz	0.47 dB 0.81 dB	
In 0.01 dB steps (0 to -1.99) dBm	100 kHz to 2.56 GHz SWR ≤ 1.5:1	0.036 dB	HP 8663A, Type-N(f)

Parameter/Range	Frequency	CMC ² (±)	Comments
RF Absolute Power – Generate ^{3,6} – (cont)			
Into 50 Ω (+16 to -119.9) dBm	100 kHz to 2.56 GHz SWR ≤ 1.5:1	1.2 dB	HP 8663A, Type-N(f)
(-120 to -129.9) dBm	100 kHz to 2.56 GHz SWR ≤ 1.5:1	3.5 dB	HP 8663A, Type-N(f)
> +10 dBm	10 MHz to 2 GHz SWR ≤ 1.6:1 (≥ 2 to ≤ 20) GHz SWR ≤ 1.6:1	1.4 dB 1.5 dB	HP 83650B, 2.4 mm(m)
> -10 dBm	10 MHz to 2 GHz SWR ≤ 1.6:1 (≥ 2 to ≤ 20) GHz SWR ≤ 1.6:1 (> 20 to ≤ 40) GHz SWR ≤ 1.8:1 (> 40 to ≤ 50) GHz SWR ≤ 2:1	0.72 dB 0.84 dB 1.1 dB 2.0 dB	HP 83650B, 2.4 mm(m)
> -60 dBm	10 MHz to 2 GHz SWR ≤ 1.6:1 (≥ 2 to ≤ 20) GHz SWR ≤ 1.6:1 (> 20 to ≤ 40) GHz SWR ≤ 1.8:1 (> 40 to ≤ 50) GHz SWR ≤ 2:1	1.1 dB 1.2 dB 1.4 dB 2.4 dB	HP 83650B, 2.4 mm(m)
≤ -60 dBm	10 MHz to 2 GHz SWR ≤ 1.6:1 (≥ 2 to ≤ 20) GHz SWR ≤ 1.6:1 (> 20 to ≤ 40) GHz SWR ≤ 1.8:1 (> 40 to ≤ 50) GHz SWR ≤ 2:1	1.7 dB 1.8 dB 2.0 dB 3.0 dB	HP 83650B, 2.4 mm(m)

Parameter/Range	Frequency	CMC ² (±)	Comments
Pulse – Generate ³			
Transition Time	(10 to 90) %	0.13 ns	HP 8133A
Width	150 ps to 10 ns	0.13 ns	HP 8133A
	10 ns to < 100 ns	(0.013 x Width) + 1.2 ns	HP 8161A
	≥ 100 μs to < 10 ms	(0.013 x Width) + 0.14 μs	HP 8161A
	≥ 10 ms to < 100 ms	(0.012 x Width) + 2.0 ns	HP 8161A
	≥ 100 ms to < 0.99 ms	(0.012 x Width) + 0.19 μs	HP 8161A
RMS Jitter – Period, Delay and Width	33 MHz to 3.0 GHz	10 ps	HP 8133A
Pulse – Measure			
RMS Jitter – Period, Delay and Width	33 MHz to 3.0 GHz	6.6 ps	HP 54124T or HP 86100
Attenuation – Generate ³			
Coaxial, 1 dB Step (0 to 11) dB			
0 dB	50 MHz to 2 GHz (2 to 4) GHz	0.027 dB 0.029 dB	HP 8494G w/ Type-N(f)
1 dB	50 MHz to 2 GHz (2 to 4) GHz	0.027 dB 0.029 dB	
2 dB	50 MHz to 2 GHz (2 to 4) GHz	0.027 dB 0.029 dB	
3 dB	50 MHz to 2 GHz (2 to 4) GHz	0.028 dB 0.029 dB	
4 dB	50 MHz to 2 GHz (2 to 4) GHz	0.028 dB 0.030 dB	
5 dB	50 MHz to 2 GHz (2 to 4) GHz	0.029 dB 0.030 dB	

Parameter/Range	Frequency	CMC ² (±)	Comments	
Attenuation – Generate ³				
Coaxial, 1 dB Step (0 to 11) dB				
6 dB	50 MHz to 2 GHz (2 to 4) GHz	0.030 dB 0.030 dB	HP 8494G w/ Type-N(f)	
7 dB	50 MHz to 2 GHz (2 to 4) GHz	0.030 dB 0.030 dB		
8 dB	50 MHz to 2 GHz (2 to 4) GHz	0.030 dB 0.030 dB		
9 dB	50 MHz to 2 GHz (2 to 4) GHz	0.031 dB 0.030 dB		
10 dB	50 MHz to 2 GHz (2 to 4) GHz	0.031 dB 0.030 dB		
11 dB	50 MHz to 2 GHz (2 to 4) GHz	0.031 dB 0.030 dB		
Coaxial, 10 dB Step				
0 dB	50 MHz to 2 GHz (2 to 4) GHz	0.027 dB 0.029 dB		
10 dB	50 MHz to 2 GHz (2 to 4) GHz	0.031 dB 0.030 dB		
20 dB	50 MHz to 2 GHz (2 to 4) GHz	0.033 dB 0.032 dB		
30 dB	50 MHz to 2 GHz (2 to 4) GHz	0.036 dB 0.035 dB		
40 dB	50 MHz to 2 GHz (2 to 4) GHz	0.054 dB 0.048 dB		
50 dB	50 MHz to 2 GHz (2 to 4) GHz	0.062 dB 0.057 dB		

Parameter/Range	Frequency	CMC ² (±)	Comments
Attenuation – Generate ³ (cont)			
Coaxial, 10 dB Step			
60 dB	50 MHz to 2 GHz (2 to 4) GHz	0.064 dB 0.058 dB	HP 8496G w/ Type-N(f)
70 dB	50 MHz to 2 GHz (2 to 4) GHz	0.070 dB 0.060 dB	
80 dB	50 MHz to 2 GHz (2 to 4) GHz	0.084 dB 0.068 dB	
90 dB	50 MHz to 2 GHz (2 to 4) GHz	0.091 dB 0.074 dB	
100 dB	50 MHz to 2 GHz (2 to 4) GHz	0.093 dB 0.076 dB	
110 dB	50 MHz to 2 GHz (2 to 4) GHz	0.11 dB 0.077 dB	
Coaxial, Fixed			
3 dB	DC to 2 GHz SWR < 1.25:1	0.028 dB	HP 8491A/B Type-N
	(2 to 4) GHz SWR < 1.2:1	0.029 dB	
	(4 to 18) GHz SWR < 1.2:1	0.053 dB	
6 dB	DC to 2 GHz SWR < 1.25:1	0.029 dB	
	(2 to 4) GHz SWR < 1.2:1	0.030 dB	
	(4 to 18) GHz SWR < 1.2:1	0.053 dB	

Parameter/Range	Frequency	CMC ² (±)	Comments
Attenuation – Generate ³ (cont)			
Coaxial, Fixed			
10 dB	DC to 2 GHz SWR < 1.25:1	0.031 dB	HP 8491A/B Type-N
	(2 to 4) GHz SWR < 1.2:1	0.030 dB	
	(4 to 18) GHz SWR < 1.2:1	0.053 dB	
20 dB	DC to 2 GHz SWR < 1.5:1	0.032 dB	
	(2 to 4) GHz SWR < 1.5:1	0.032 dB	
	(4 to 18) GHz SWR < 1.5:1	0.053 dB	

Parameter/Range	CMC ² (±)	Comments
Reflection S ₁₁ / S ₂₂ – Measure ³		
30 kHz to 1.3 GHz (0 to 1.0) lin	(± 0.0044 to ± 0.022) lin (± 180 to ± 1.3) deg	Network analyzer HP8753ES Type-N precision cal kit HP85032B APC 7mm precision cal kit HP85031B
300 kHz to 1.3 GHz (0 to 1.0) lin	(± 0.0071 to ± 0.035) lin (± 180 to ± 2.0) deg	
300 kHz to 1.3 GHz (0 to 1.0) lin	(± 0.0038 to ± 0.017) lin (± 180 to ± 0.96) deg	Network analyzer HP8753ES Type N precision cal kit HP85032B
(1.3 to 3) GHz (0 to 1.0) lin	(± 0.0051 to ± 0.026) lin (± 180 to ± 1.5) deg	
(3 to 6) GHz (0 to 1.0) lin	(± 0.011 to ± 0.050) lin (± 180 to ± 3.0) deg	

Parameter/Range	CMC ² (±)	Comments
Reflection S ₁₁ / S ₂₂ – Measure ³ (cont) 300 kHz to 1.3 GHz (0 to 1.0) lin (1.3 to 3) GHz (0 to 1.0) lin (3 to 6) GHz (0 to 1.0) lin	(± 0.0019 to ± 0.0072) lin (± 180 to ± 0.42) deg (± 0.0036 to ± 0.011) lin (± 180 to ± 0.59) deg (± 0.0057 to ± 0.018) lin (± 180 to ± 1.1) deg	Network analyzer HP8753ES APC 7mm precision cal kit HP85031B
Transmission S ₁₂ / S ₂₁ – Measure ³ 30 kHz to 1.3 GHz (0 to 20) dB (20 to 40) dB (40 to 60) dB (1.3 to 3) GHz (0 to 20) dB (20 to 40) dB (40 to 60) dB (3 to 6) GHz (0 to 20) dB (20 to 40) dB (40 to 60) dB	(± 0.041 to ± 0.063) dB (± 0.42 to ± 0.46) deg (± 0.063 to ± 0.088) dB (± 0.46 to ± 0.62) deg (± 0.088 to ± 0.29) dB (± 0.62 to ± 2.0) deg (± 0.056 to ± 0.076) dB (± 0.79 to ± 1.3) deg (± 0.076 to ± 0.11) dB (± 1.3 to ± 1.4) deg (± 0.11 to ± 0.32) dB (± 1.4 to ± 2.2) deg (± 0.094 to ± 0.13) dB (± 1.2 to ± 1.6) deg (± 0.13 to ± 0.16) dB (± 1.6 to ± 1.7) deg (± 0.16 to ± 0.51) dB (± 1.7 to ± 3.6) deg	Network Analyzer HP8753ES Type-N Precision Cal Kit HP85032B

III. Time and Frequency

Parameter/Equipment	Range	CMC ² (±)	Comments
Time Interval – Measure ³	10 ns to 10 s	2.9 ns	HP 5334B/Datum 8040
Frequency – Measuring Equipment ³	5 MHz, 10 MHz	10 pHz/Hz	Datum 8040
Frequency – Measure ³	1 Hz to 40 GHz	50 pHz/Hz	HP 53132A HP 5352B

¹ This laboratory offers calibration service only for Motorola in Plantation Florida.

² Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ Field calibration service is available for this calibration and this laboratory meets A2LA R104 – *General Requirements: Accreditation of Field Testing and Field Calibration Laboratories* for these calibrations. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

⁴ Based on using the standard at the temperature the Fluke 5720A/5725A was calibrated ($t_{cal} \pm 5 \text{ }^\circ\text{C}$) and assuming the instrument is zeroed at least every seven days or when the ambient temperature changes more than $5 \text{ }^\circ\text{C}$. For resistance a zero calibration is performed at least every 12 hours within $\pm 1 \text{ }^\circ\text{C}$ of use. CMC is based upon 1-year specifications and is read as a portion or percent output plus floor specification. The floor specification is expressed as a fixed value or a portion or percent of the range.

⁵ Based on using the HP 3458A at the temperature (t_{cal}) it was calibrated $\pm 5 \text{ }^\circ\text{C}$ and an auto-calibration (ACAL) was performed within the previous 24 hours ($\pm 1 \text{ }^\circ\text{C}$ of ambient temperature). CMC is based upon 1-year specifications and is read as a portion or percent output plus floor specification. The floor specification is expressed as a fixed value or a portion or percent of the range.

⁶ CMC does not include the mismatch.

⁷ Ranges are based upon the system combination used:

Instrument/System	Ranges
HP 8902A	2.5 MHz to 1.3 GHz Range 1 & 2 – SWR 1.18:1 Range 3 – SWR 1.4:1
HP 8902A w/ HP 11722A	2.5 MHz to 1.3 GHz Range 1 & 2 – SWR 1.33:1 Range 3 – SWR 1.5:1
HP 8902A w/ HP 11792A or HP 11793A	(50 to 1300) MHz SWR 1.15:1 (1.3 to 18) GHz SWR 1.25:1 (18 to 26.5) GHz SWR 1.4:1

⁸ CMCs are based upon the AM depths. For depths between 0 % and 9.99 %, the digit uncertainty (resolution) is 0.01 %. For depths between 10 % and 99.9 % FS, the digit uncertainty (resolution) is 0.1 %.

⁹ CMCs are based upon the peak phase deviations. For deviations between 0 and 3.999, the digit uncertainty (resolution) is 1 Hz. For deviations between 0 and 39.99, the digit uncertainty (resolution) is 10 Hz. For deviations between 40 and 400, the digit uncertainty (resolution) is 100 Hz.

¹⁰ *AM* is the amplitude modulation; *FM* is the frequency modulation; and *TI* is the time interval.

¹¹ All readings stated in percent mean percent of reading



World Class Accreditation

The American Association for Laboratory Accreditation

Accredited Laboratory

A2LA has accredited

AGILENT TECHNOLOGIES PLANTATION LOCAL CALIBRATION CENTER

Plantation, FL

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 *General Requirements for the Competence of Testing and Calibration Laboratories*. This laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and the requirements of ANSI/NCSL Z540.3-2006 and any additional program requirements in the field of calibration. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated 8 January 2009).



Presented this 24th day of March 2011.

A handwritten signature in black ink, appearing to read "Peter Abney", written over a horizontal line.

President & CEO
For the Accreditation Council
Certificate Number 2600.01
Valid to July 31, 2013

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.