



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

PRATT AND WHITNEY MEASUREMENT SYSTEMS, INC  
 66 Douglas Street  
 Bloomfield, CT 06002  
 Matthew Noonan Phone: 800 371 7174 ext. 440

CALIBRATION

Valid to: May 31, 2012

Certificate Number: 2629.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations and dimensional inspections<sup>1</sup>:

I. Dimensional

Parameter/Equipment	Range	CMC <sup>2,4</sup> (±)	Comments
Laseruler	(0.5 to 7.5) in	(2.3 + 1.0L) µin	Gage blocks
Labmaster Universal	(0.5 to 6.5) in	(2.3 + 1.0L) µin	Gage blocks
Labmaster Universal Model 1000A	(0.0 to 80) in	(2.3 + 1.0L) µin	Gage blocks, force gages
Labmaster Standard	(0.5 to 7.5) in	(2.3 + 1.0L) µin	Gage blocks
Labmicrometer	(0.5 to 7.5) in	(2.3 + 1.0L) µin	Gage blocks
Universal Supermicrometer <sup>3</sup>	(0 to 2) in	5.8 µin	Gage blocks

Parameter/Equipment	Range	CMC <sup>2,4</sup> (±)	Comments
Supermicrometer Model C <sup>3</sup>	(0 to 1) in	15 μin	Gage blocks
Supermicrometer Model B <sup>3</sup>	(0 to 1) in	18 μin	Gage blocks
Digital Measuring Machine –  Lead Screw  Machine Button Bar	(0 to 1) in  (1 to 20) in	15 μin  (49 + 0.8L) μin	Gage blocks
Standard Measuring Machine –  Lead Screw  Machine Button Bar	(0 to 1) in  (1 to 20) in	18 μin  (49 + 0.8L) μin	Gage blocks
Plain Cylindrical Ring Gages	(0.020 to 1) in (1 to 14) in	6.5 μin (3.5 + 0.8L) μin	Labmaster Universal

## II. Mechanical

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Labmaster Universal Force System <sup>3</sup> –  Fixed Points	2 ozf 8 ozf 16 ozf 24 ozf 32 ozf 40 ozf	0.17 ozf 1.4 ozf 1.5 ozf 1.7 ozf 1.8 ozf 2.0 ozf	Force gage

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Labmicrometer Force System <sup>3</sup> –  Fixed Points	-8 ozf -2 ozf 2 ozf 8 ozf 16 ozf 24 ozf 32 ozf 40 ozf	0.22 ozf 0.10 ozf 0.10 ozf 0.22 ozf 0.41 ozf 0.60 ozf 0.80 ozf 1.0 ozf	Force gage
Laser Measuring Machine Force System <sup>3</sup> –  Fixed Points	2 ozf 8 ozf 16 ozf 24 ozf 32 ozf 40 ozf	0.17 ozf 1.4 ozf 1.5 ozf 1.7 ozf 1.8 ozf 2.0 ozf	Force gages

<sup>1</sup> This laboratory offers commercial and field calibration service.

<sup>2</sup> Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

<sup>3</sup> Field calibration service is available for this calibration and this laboratory meets A2LA R104 – *General Requirements: Accreditation of Field Testing and Field Calibration Laboratories* for these calibrations. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

<sup>4</sup> In the statement of the CMC,  $L$  is the numerical value of the nominal length of the device measured in inches.



The American Association for Laboratory Accreditation

World Class Accreditation

# Accredited Laboratory

A2LA has accredited

**PRATT & WHITNEY  
MEASUREMENT SYSTEMS, INC.**

*Bloomfield, CT*

for technical competence in the field of

## Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 *General Requirements for the Competence of Testing and Calibration Laboratories*. This laboratory also meets any additional program requirements in the field of calibration. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated 8 January 2009).



Presented this 4<sup>th</sup> day of June 2010.

  
\_\_\_\_\_  
Peter M. Meyer

President & CEO  
For the Accreditation Council  
Certificate Number 2629.01  
Valid to May 31, 2012  
Revised April 27, 2012

*For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.*