



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005
& ANSI/NCSL Z540-1-1994 & ANSI/NCSL Z540.3-2006

AGILENT TECHNOLOGIES CANADA SERVICE CENTER
2250 boul ALFRED NOBEL
SAINT-LAURENT, QC
CANADA, H4S 2C9
Jim Nagy Phone: 514 832 2836

CALIBRATION

Valid To: January 31, 2012

Certificate Number: 2645.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations¹:

I. Electrical – DC/Low Frequency

Parameter/Equipment	Range	CMC ^{2,4,5,12} (±)	Comments
DC Voltage – Generate ³	(0 to 0.22) V (0.22 to 2.2) V (2.2 to 11) V (11 to 22) V (22 to 220) V (220 to 1100) V	3.8 µV/V + 2.5 µV 4.9 µV/V + 1.9 µV 3.3 µV/V + 4.7 µV 2.4 µV/V + 48 µV 5.4 µV/V + 40 µV 6.9 µV/V + 520 µV	Fluke 5720A/5725A
DC Voltage – Measure ³	(0 to 0.1) V (0.1 to 1) V (1 to 10) V (10 to 100) V (100 to 1000) V	5.6 µV/V + 1.5 µV 5.2 µV/V + 1.2 µV 4.7 µV/V + 2.5 µV 6.6 µV/V + 45 µV 19 µV/V + 160 µV	HP 3458A/100 PLC option 002
Shunt 15 A 0.1 Ω, 25 W	Dissipated power: 0.3 Amps to full power	0.014 %	Guildline 9230-15
Shunt 100A 0.01 Ω, 100 W	≤ 25W	0.014 %	Guildline 9230-100

Parameter/Equipment	Range	CMC ^{2,4,5} (\pm)	Comments
DC Current – Generate ³	(0 to 220) μ A 220 μ A to 2.2 mA (2.2 to 22) mA (22 to 100) mA (100 to 220) mA	36 μ A/A + 0.009 μ A 33 μ A/A + 0.008 μ A 34 μ A/A + 0.040 μ A 42 μ A/A + 0.68 μ A 51 μ A/A + 0.00 μ A	Fluke 5720A
	220 mA to 1 A (1 to 2.2) A (2.2 to 11) A	76 μ A/A + 12 μ A 0.015 % - 66 μ A 0.030 % + 400 μ A	Fluke 5720A/5725A
DC Current – Measure ³	(10 to 100) μ A 100 μ A to 1 mA (1 to 10) mA (10 to 100) mA 100 mA to 1.1 A	20 μ A/A + 0 μ A 20 μ A/A + 0.1 μ A 20 μ A/A + 0.06 μ A 35 μ A/A + 0.60 μ A 0.011 % + 11 μ A	HP 3458A
Resistance – Generate, Fixed Points ³	0 Ω 1 Ω 1.9 Ω 10 Ω 19 Ω 100 Ω 190 Ω 1 k Ω 1.9 k Ω 10 k Ω 19 k Ω 100 k Ω 190 k Ω 1 M Ω 1.9 M Ω 10 M Ω 19 M Ω 100 M Ω	0.25 m Ω 0.27 m Ω 0.31 m Ω 0.34 m Ω 2.5 m Ω 2.7 m Ω 3.3 m Ω 9.3 m Ω 31 m Ω 93 m Ω 0.19 Ω 1.2 Ω 2.2 Ω 20 Ω 42 Ω 400 Ω 1.5 k Ω 12 k Ω	Fluke 5720A
	1 Ω 10 k Ω	8 $\mu\Omega/\Omega$ 4 $\mu\Omega/\Omega$ (trend \pm 1 $\mu\Omega/\Omega$)	Fluke 742-1 Fluke 742-1
Resistance – Measure ³	(0 to 10) Ω (10 to 100) Ω 100 to 1 k Ω (1 to 10) k Ω (10 to 100) k Ω 100 k Ω to 1 M Ω (1 to 10) M Ω (10 to 100) M Ω 100 M Ω to 1 G Ω	18 $\mu\Omega/\Omega$ + 74 $\mu\Omega$ 13 $\mu\Omega/\Omega$ + 740 $\mu\Omega$ 11 $\mu\Omega/\Omega$ + 7.6 $\mu\Omega$ 11 $\mu\Omega/\Omega$ + 74 $\mu\Omega$ 11 $\mu\Omega/\Omega$ + 740 $\mu\Omega$ 15 $\mu\Omega/\Omega$ + 2.4 $\mu\Omega$ 53 $\mu\Omega/\Omega$ + 110 $\mu\Omega$ 0.05 % + 4.3 k Ω 0.5 % + 290 k Ω	HP 3458A

Parameter/Range	Frequency	CMC ^{2,4,5} (±)	Comments
AC Current – Generate ³			
(0 to 220) μA	(10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.016 % + 0.063 μA 89 μA/A + 0.062 μA 60 μA/A + 0.062 μA 0.018 % + 0.062 μA 0.10 % + 0.09 μA	Fluke 5720A
220 μA to 2.2 mA	(10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.027 % + 0.056 μA 0.020 % + 0.051 μA 0.016 % + 0.052 μA 0.023 % + 0.12 μA 0.11 % + 0.67 μA	
(2.2 to 22) mA	(10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.027 % + 0.56 μA 0.021 % + 0.51 μA 0.016 % + 0.52 μA 0.023 % + 0.71 μA 0.11 % + 5.1 μA	
(22 to 220) mA	(10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.028 % + 3.9 μA 0.021 % + 3.0 μA 0.017 % + 2.2 μA 0.024 % + 3.1 μA 0.11 % + 0.01 mA	
220 mA to 2.2 A	20 Hz to 1 kHz 220 mA to 1.0 A >1 A to 2.2 A (1 to 5) kHz 220 mA to 1.0 A >1 A to 2.2 A (5 to 10) kHz	0.030 % + 32 μA 0.035 % + 26 μA 0.044 % + 83 μA 0.055 % + 72 μA 0.67 % + 0.072 mA	
(2.2 to 11) A	40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.044 % + 130 μA 0.088 % + 290 μA 0.31 % + 0.64 mA	
AC Current – Measure ³			
(0 to 100) μA	(10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz	0.4 % + 31 nA 0.15 % + 31 nA 0.06 % + 31 nA	HP 3458A
(0.1 to 1) mA	(10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz	0.4 % + 310 nA 0.15 % + 210 nA 0.06 % + 210 nA	

Parameter/Range	Frequency	CMC ^{2,4,5} (±)	Comments
AC Current – Measure (cont) ³			
(1 to 10) mA	(10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz	0.4 % + 3.1 µA 0.15 % + 2.1 µA 0.06 % + 2.1 µA	HP 3458A
(10 to 100) mA	(10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz	0.4 % + 31 µA 0.15 % + 21 µA 0.06 % + 21 µA	
100 mA to 1.05 A	(10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz	0.4 % + 220 µA 0.16 % + 220 µA 0.08 % + 220 µA 0.1 % + 220 µA	
AC Voltage – Generate ³			
(0 to 22) mV	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.027 % + 4.5 µV 0.013 % + 4.5 µV 0.011 % + 4.5 µV 0.027 % + 4.5 µV 0.054 % + 5.3 µV 0.12 % + 10 µV 0.17 % + 24 µV 0.30 % + 24 µV	Fluke 5720A
(22 to 220) mV	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.025 % + 12 µV 95 µV/V + 7.3 µV 86 µV/V + 7.3 µV 0.018 % + 9.4 µV 0.049 % + 16 µV 0.088 % + 20 µV 0.14 % + 0.033 mV 0.27 % + 0.048 mV	
220 mV to 2.2 V	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.025 % + 39 µV 99 µV/V + 15 µV 63 µV/V + 5.9 µV 86 µV/V + 8.3 µV 0.011 % + 30 µV 0.041 % + 78 µV 0.99 % + 200 µV 0.16 % + 0.32 mV	

Parameter/Range	Frequency	CMC ^{2,4,5,12} (±)	Comments
AC Voltage – Generate (cont) ³			
(2.2 to 22) V	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.026 % + 420 μV 0.011 % + 150 μV 67 μV/V + 44 μV 91 μV/V + 90 μV 0.011 % + 210 μV 0.029 % + 650 μV 0.11 % + 2.0 mV 0.16 % + 3.3 mV	Fluke 5720A
(22 to 220) V	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (>22 to 100) Vrms (>100 to 220) Vrms (50 to 100) kHz (>22 to 100) Vrms (>100 to 220) Vrms (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.024 % + 3.9 mV 0.010 % + 1.5 mV 69 μV/V + 430 μV 98 μV/V + 910 μV 0.017 % + 560 μV 0.016 % + 2.4 mV 0.021 % + 1.9 mV 0.087 % + 16 mV 0.43 % + 39 mV 0.79 % + 79 mV	
(0 to 250) V max output, 1100 V range	(15 to 50) Hz 50 Hz to 1 kHz	0.031 % + 17 mV 91 μV/V + 2.9 mV	Fluke 5720A/ 5725A
(0 to 250) V max output, 1100 V range	40 Hz to 20 kHz (20 to 30) kHz	91 μV/V + 2.9 mV 0.051 % + 9.6 mV	
(0 to 750) V	(30 to 50) kHz (50 to 100) kHz	0.052 % + 8.6 mV 0.19 % + 37 mV	
AC Voltage – Measure ^{3,12}			
(0 to 10) mV	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 4) MHz (4 to 8) MHz	0.03 % + 3.1 μV 0.02 % + 1.2 μV 0.03 % + 1.7 μV 0.1 % + 1.6 μV 0.5 % + 1.3 μV 4 % + 2.1 μV 1.2 % + 6.6 μV 7 % + 7.5 μV 20 % + 8.2 μV	HP 3458A

Parameter/Range	Frequency	CMC ^{2,5} (\pm)	Comments
AC Voltage – Measure (cont) ^{3,12}			
(10 to 100) mV	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz (2 to 4) MHz (4 to 8) MHz (8 to 10) MHz	70 μ V/V + 4.1 μ V 70 μ V/V + 2.1 μ V 0.014 % + 2.3 μ V 0.03 % + 2.6 μ V 0.08 % + 2.3 μ V 0.3 % + 15 μ V 1 % + 28 μ V 1.5 % + 20 μ V 4 % + 74 μ V 4 % + 83 μ V 15 % + 110 μ V	HP 3458A
100 mV to 1 V	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz (2 to 4) MHz (4 to 8) MHz (8 to 10) MHz	70 μ V/V + 41 μ V 70 μ V/V + 21 μ V 0.014 % + 22 μ V 0.03 % + 22 μ V 0.08 % + 22 μ V 0.3 % + 120 μ V 1 % + 300 μ V 1.5 % + 210 μ V 4 % + 730 μ V 4 % + 830 μ V 15 % + 1 mV	
(1 to 10) V	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz (2 to 4) MHz (4 to 8) MHz (8 to 10) MHz	70 μ V/V + 420 μ V 70 μ V/V + 220 μ V 0.014 % + 240 μ V 0.03 % + 250 μ V 0.08 % + 220 μ V 0.3 % + 1.1 mV 1 % + 1.1 mV 1.5 % + 1.1 mV 4 % + 7.1 mV 4 % + 8.1 mV 15 % + 11 mV	
(10 to 100) V	(1 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz	0.02 % + 4.1 mV 0.02 % + 2.6 mV 0.035 % + 2.4 mV 0.12 % + 2.1 mV 0.4 % + 11 mV 1.5 % + 40 mV	
(100 to 750) V	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz	0.04 % + 31 mV 0.04 % + 16 mV 0.06 % + 16 mV 0.12 % + 16 mV 0.3 % + 15 mV	

Parameter/Range	Frequency	CMC ^{2, 4, 12} (±)	Comments
AC Voltage Flatness – Generate ³			
0.3 mV to 3.5 V	(10 to 30) Hz 30 Hz to 120 kHz	0.27 % 0.14 %	Fluke 5720A or 5700A-03 (referenced to 1 kHz)
(0.3 to 1.1) mV	120 kHz to 2 MHz (2 to 10) MHz (10 to 20) MHz (20 to 30) MHz	0.46 % 0.62 % 0.80 % 2.4 %	
(0.3 to 3) mV	120 kHz to 2 MHz (2 to 10) MHz (10 to 20) MHz (20 to 30) MHz	0.22 % 0.37 % 0.55 % 1.4 %	
3 mV to 3.5 V	120 kHz to 2 MHz (2 to 10) MHz (10 to 20) MHz (20 to 30) MHz	0.12 % 0.21 % 0.38 % 0.86 %	
AC Voltage Flatness – Measure ³			
Up to 3 V	10 Hz 100 Hz (10, 30) kHz 100 kHz 300 kHz 1 MHz 3 MHz 8 MHz 10 MHz 20 MHz 30 MHz 50 MHz 70 MHz 80 MHz 100 MHz	0.02 % + 6.9 μV/V 80 μV + 5.5 μV/V 80 μV + 3.2 μV/V 0.01 % + 8.0 μV/V 0.01 % + 5.2 μV/V 0.01 % + 6.5 μV/V 0.13 % + 59 μV/V 0.13 % + 110 μV/V 0.13 % + 91 μV/V 0.25 % + 210 μV/V 0.25 % + 240 μV/V 0.61 % + 340 μV/V 0.9 % + 240 μV/V 1.1 % + 790 μV/V 1.3 % + 940 μV/V	By comparison to 11049A, 11050A, 11051A thermal voltage converters

Parameter/Range	Frequency	CMC ^{2, 4, 12} (±)	Comments
Capacitance Generate (cont) ³ – Algorithmic Derivation 1000 pF	1 MHz 2 MHz 3 MHz 4 MHz 5 MHz 10 MHz 13 MHz	0.05 % 0.06 % 0.10 % 0.15 % 0.20 % 0.50 % 0.70 %	16380A/16380C standard air capacitor set, BNC 4 terminal pair
Substitution Method (0.01, 0.1, 1) μF	120 Hz to 10 kHz	0.025 % 0.04 %	
(0.01, 0.1) μF 1 μF	100 kHz	0.05 % 0.1 %	
Direct Measure 3.3 to 10.9999 nF 11 to 32.9999 nF 0.33 to 109.999 μF 110 to 329.999 μF 3.3 to 10.9999 μF	10 to 1000 Hz 10 to 1000 Hz 10 to 600 Hz 10 to 300Hz 10 to 150Hz	0.4 % 0.4 % 0.4 % 0.4 % 0.4 %	Fluke 5720A
Voltage Ratio – Generate Decades ³ 1 and 2 All Other Decades	1 kHz	1.1 x 10 ⁻⁶ input 0.51 x 10 ⁻⁶ input	DT-72 ratio transformer

Parameter/Equipment	Range	CMC ² (±)	Comments
Electrical Calibration of Thermocouple Indicators ³ – Type J	(-210 to -100) °C (-100 to -30) °C (-30 to 150) °C (150 to 760) °C (760 to 1200) °C	0.36 °C 0.23 °C 0.21 °C 0.24 °C 0.32 °C	Fluke 5520

II. Electrical – RF/Microwave

Parameter/Equipment/Range	CMC ^{2, 11, 12} (±)	Comments
<p>Amplitude Modulation – Measure^{3, 8}</p> <p>(0.15 to 10) MHz (5% to < 10%) AM (10% to 99%) AM</p> <p>(5% to < 10%) AM (10% to 99%) AM</p> <p>(10 to 1300) MHz (5% to <10%) AM (10% to 99%) AM</p> <p>(5% to <10%) AM (10% to 99%) AM</p> <p>1300 MHz to 26.5 GHz (5% to <10%) AM (10% to 99%) AM</p> <p>10 MHz to 26.5 GHz (5% to <10%) AM (10% to 99%) AM</p>	<p>0.025AM + 0.030 % 0.025AM + 0.14 %</p> <p>0.038AM + 0.030 % 0.038AM + 0.13 %</p> <p>0.012AM + 0.033 % 0.012AM + 0.17 %</p> <p>0.038AM + 0.030 % 0.037AM + 0.16 %</p> <p>0.019AM + 0.029 % 0.019AM + 0.14 %</p> <p>0.038AM + 0.030 % 0.038AM + 0.11 %</p>	<p>HP 8902 measuring receiver:</p> <p>rate: 50 Hz to 10 kHz, depth: 5% to 99%</p> <p>rate: 20 Hz to 10 kHz, depth: to 99%</p> <p>rate: 50 Hz to 10 kHz, depth: 5% to 99%</p> <p>rate: 20 Hz to 10 kHz, depth: to 99%</p> <p>rate: 50 Hz to 10 kHz, depth: 5% to 99%</p> <p>rate: 20 Hz to 10 kHz, depth: to 99%</p>
<p>Frequency Modulation – Measure^{3, 9}</p> <p>(0.25 to 10) MHz (0 to <4) kHz Peak FM (≥ 4 to < 40) kHz Peak FM</p> <p>(10 to 1300) MHz (0 to <4) kHz Peak FM (≥ 4 to < 40) kHz Peak FM (≥ 40 to < 400) kHz Peak FM</p> <p>(10 to 1300) MHz (0 to <4) kHz Peak FM (≥ 4 to < 40) kHz Peak FM (≥ 40 to < 400) kHz Peak FM</p>	<p>0.024FM + 2.6 Hz Pk 0.024FM + 10 Hz Pk</p> <p>0.012FM + 2.6 Hz Pk 0.012FM + 12 Hz Pk 0.012FM + 110 Hz Pk</p> <p>0.059FM + 2.8 Hz Pk 0.059FM + 13 Hz Pk 0.059FM + 110 Hz Pk</p>	<p>HP 8902 measuring receiver: 40.0 to 400.0: 1 digit = 100 Hz 4.00 to 39.99: 1 digit = 10 Hz 0 to 3.999: 1 digit = 1 Hz</p> <p>rate: 20 Hz to 10 kHz, ≤40 kHz peak</p> <p>rate: 50 Hz to 100 kHz, ≤400 kHz peak</p> <p>rate: 20 Hz to 200 kHz, ≤400 kHz peak</p>

Parameter/Equipment/Range	CMC ^{2, 11} (±)	Comments
<p>Frequency Modulation – Measure (cont)^{3, 9}</p> <p>10 MHz to 26.5 GHz (10 to 1300) MHz (0 to <4) kHz Peak FM (≥ 4 to < 40) kHz Peak FM (≥ 40 to < 400) kHz Peak FM</p> <p>(>1.3 to 6.2) GHz (0 to <4) kHz Peak FM (≥ 4 to < 40) kHz Peak FM (≥ 40 to < 400) kHz Peak FM</p> <p>(>6.2 to 12.4) GHz (0 to <4) kHz Peak FM (≥ 4 to < 40) kHz Peak FM (≥ 40 to < 400) kHz Peak FM</p> <p>(>12.4 to 18.6) GHz (0 to <4) kHz Peak FM (≥ 4 to < 40) kHz Peak FM (≥ 40 to < 400) kHz Peak FM</p> <p>(>18.6 to 26.5) GHz (0 to <4) kHz Peak FM (≥ 4 to < 40) kHz Peak FM (≥ 40 to < 400) kHz Peak FM</p> <p>10 MHz to 26.5 GHz (10 to 1300) MHz (0 to <4) kHz Peak FM (≥ 4 to < 40) kHz Peak FM (≥ 40 to < 400) kHz Peak FM</p> <p>(>1.3 to 6.2) GHz (0 to <4) kHz Peak FM (≥ 4 to < 40) kHz Peak FM (≥ 40 to < 400) kHz Peak FM</p> <p>(>6.2 to 12.4) GHz (0 to <4) kHz Peak FM (≥ 4 to < 40) kHz Peak FM (≥ 40 to < 400) kHz Peak FM</p> <p>(>12.4 to 18.6) GHz (0 to <4) kHz Peak FM (≥ 4 to < 40) kHz Peak FM (≥ 40 to < 400) kHz Peak FM</p>	<p>0.012FM + 2.4 Hz Pk 0.012FM + 12 Hz Pk 0.012FM + 100 Hz Pk</p> <p>0.0099FM + 10 Hz Pk 0.012FM + 12 Hz Pk 0.012FM + 100 Hz Pk</p> <p>0.0075FM + 24 Hz Pk 0.012FM + 12 Hz Pk 0.012FM + 85 Hz Pk</p> <p>0.0049FM + 52 Hz Pk 0.011FM + 36 Hz Pk 0.012FM + 110 Hz Pk</p> <p>0.0035FM + 80 Hz Pk 0.011FM + 46 Hz Pk 0.012FM + 100 Hz Pk</p> <p>0.059FM + 2.8 Hz Pk 0.059FM + 14 Hz Pk 0.059FM + 120 Hz Pk</p> <p>0.058FM + 5.2 Hz Pk 0.059FM + 14 Hz Pk 0.059FM + 120 Hz Pk</p> <p>0.055FM + 15 Hz Pk 0.059FM + 14 Hz Pk 0.059FM + 120 Hz Pk</p> <p>0.050FM + 37 Hz Pk 0.059FM + 15 Hz Pk 0.059FM + 120 Hz Pk</p>	<p>HP 8902 measuring receiver: 11793A DownConverter, Gen. Ext L.O.</p> <p>rate: 50 Hz to 100 kHz, ≤400 kHz peak</p> <p>rate: 20 Hz to 200 kHz, ≤400 kHz peak</p>

Peter Abney

Parameter/Equipment/Range	CMC ^{2, 11} (±)	Comments
Frequency Modulation – Measure (cont) ³ (>18.6 to 26.5) GHz (0 to <4) kHz Peak FM (≥ 4 to < 40) kHz Peak FM (≥ 40 to < 400) kHz Peak FM	 $0.045FM + 60 \text{ Hz Pk}$ $0.059FM + 16 \text{ Hz Pk}$ $0.059FM + 120 \text{ Hz Pk}$	HP 8902 measuring receiver: 11793A DownConverter, Gen. Ext L.O. rate: 20 Hz to 200 kHz, ≤400 kHz peak

Parameter/Range	Frequency	CMC ^{2, 12} (±)	Comments
Digital Modulation – Measure ³ Carrier: 2 MHz to 2.65 GHz Error Vector Magnitude for Modulation Types: MSK, GMSK, BPSK, DQPSK, $\pi/4$ DQPSK, 8PSK, 16QAM and 32QAM, QPSK Phase Error for Modulation Types: MSK, GMSK, BPSK, DQPSK, $n/4$ DQPSK, 8PSK, 16QAM and 32QAM, QPSK Error Vector Magnitude for FSK Modulation	 Mod Frequency Span: (1 to 100) kHz (0.1 to 1) MHz 1 MHz to 2.65 GHz (1 to 100) kHz (0.1 to 1) MHz 1 MHz to 2.65 GHz Mod Frequency: 3.2 kHz 1.152 kHz	 0.3 % rms 0.51 % rms 0.93 % rms 0.18° rms 0.34° rms 0.57° rms 0.54 % rms 1.5 % rms	 HP 89441A Vector Signal Analyzer

Parameter/Range	Frequency	CMC ² (±)	Comments
RF Absolute Power – Measure ^{3,6}			
1 mW, Type-N(f), 50 Ω	50 MHz	0.015 dB (3.4 μW)	HP 432A w/ HP 478A-H76
(+20 to -30) dBm, 75 Ω	(100 to 600) kHz SWR ≤ 1.8:1	0.057 dB	HP 436A, HP 438A, E418A/B or E4419A/B w/ HP 8483A, Type- N(m)
	600 kHz to 2 GHz SWR ≤ 1.18:1	0.059 dB	
(+20 to -30) dBm, 50 Ω	(100 to 300) kHz SWR ≤ 1.6:1	0.05 dB	HP 436A, HP 438A, E418A/B or E4419A/B w/ HP 8482A, Type- N(m)
	300 kHz to 1 MHz SWR ≤ 1.2:1	0.049 dB	
	1 MHz to 2 GHz SWR ≤ 1.1:1	0.051 dB	
	(2 to 4.2) GHz SWR ≤ 1.3:1	0.050 dB	
(-20 to -70) dBm, 50 Ω	(10 to 30) MHz SWR ≤ 1.4:1	0.074 dB	HP 436A, HP 438A, E418A/B or E4419A/B w/ HP 8481D, Type- N(m)
	30 MHz to 4 GHz SWR ≤ 1.15:1	0.074 dB	
	(4 to 10) GHz SWR ≤ 1.21:1	0.077 dB	
	(10 to 15) GHz SWR ≤ 1.3:1	0.10 dB	
	(15 to 18) GHz SWR ≤ 1.35:1	0.11 dB	

Parameter/Range	Frequency	CMC ² (±)	Comments		
RF Absolute Power – Measure (cont) ^{3, 6}	(+20 to -30) dBm, 50 Ω	(50 to 100) MHz SWR ≤ 1.15:1	0.060 dB	HP 436A, HP 438A, E418A/B or E4419A/B w/ HP 8487A, 2.4 mm(m)	
		(0.1 to 2) GHz SWR ≤ 1.1:1	0.075 dB		
		(2 to 12.4) GHz SWR ≤ 1.15:1	0.062 dB		
		(12.4 to 18) GHz SWR ≤ 1.2:1	0.065 dB		
		(18 to 26.5) GHz SWR ≤ 1.25:1	0.099 dB		
		(26.5 to 40) GHz SWR ≤ 1.4:1	0.097 dB		
		(40 to 50) GHz SWR ≤ 1.5:1	0.13 dB		
		(-20 to -70) dBm, 50 Ω	(50 to 100) MHz SWR ≤ 1.19:1	0.054 dB	HP 436A, HP 438A, E418A/B or E4419A/B w/ HP 8487D, 2.4 mm(m)
		(0.1 to 2) GHz SWR ≤ 1.15:1	0.054 dB		
		(2 to 12.4) GHz SWR ≤ 1.2:1	0.058 dB		
		(12.4 to 18) GHz SWR ≤ 1.29:1	0.068 dB		
		(18 to 34) GHz SWR ≤ 1.37:1	0.095 dB		
		(34 to 40) GHz SWR ≤ 1.61:1	0.11 dB		
	(40 to 50) GHz SWR ≤ 1.86:1	0.19 dB			
	(+30 to -20) dBm, 50 Ω	100 kHz to 2.6 GHz SWR ≤ 1.15:1	0.071 dB	HP 8902A w/ HP 11722A, Type-N(m)	
	(+30 to -20) dBm, 50 Ω	(50 to 1300) MHz SWR ≤ 1.15:1	0.071 dB		
	(1.3 to 18) GHz SWR ≤ 1.25:1	0.081 dB			
	(18 to 26.5) GHz SWR ≤ 1.4:1	0.092 dB	HP 8902A w/ HP 11792A, APC 3.5 mm(m)		

Parameter/Range	Frequency	CMC ² (±)	Comments
Tuned RF Power³ –			
Absolute – Measure⁷			
(≤ +10 to ≥ -22) dBm	2.5 MHz to 26.5 GHz	0.17 dB	HP 8902A with HP 11722A or with HP 11792A and HP 11793A
(< -22 to ≥ -42) dBm		0.18 dB	
(< -42 to ≥ -50) dBm		0.20 dB	
(< -50 to ≥ -60) dBm		0.21 dB	
(< -60 to ≥ -72) dBm		0.22 dB	
(< -72 to ≥ -80) dBm		0.23 dB	
(< -80 to ≥ -92) dBm		0.24 dB	
(< -92 to ≥ -102) dBm		0.27 dB	
(< -102 to ≥ -110) dBm		0.28 dB	
(< -110 to ≥ -120) dBm		0.31 dB	
(< -120 to ≥ -127) dBm		0.34 dB	
Relative – Measure			
(≤ +10 to ≥ +2) dBm	2.5 MHz to 26.5 GHz	0.081 dB	HP 8902A with HP 11722A or with HP 11792A and HP 11793A
(< +2 to ≥ -12) dBm		0.071 dB	
(< -12 to ≥ -22) dBm		0.081 dB	
(< -22 to ≥ -31) dBm		0.088 dB	
(< -31 to ≥ -40) dBm		0.095 dB	
(< -40 to ≥ -50) dBm		0.12 dB	
(< -50 to ≥ -61) dBm		0.15 dB	
(< -61 to ≥ -71) dBm		0.16 dB	
(< -71 to ≥ -80) dBm		0.17 dB	
(< -80 to ≥ -90) dBm		0.19 dB	
(< -90 to ≥ -100) dBm		0.22 dB	
(< -100 to ≥ -110) dBm		0.23 dB	
(< -110 to ≥ -120) dBm		0.27 dB	
(< -120 to ≥ -127) dBm		0.30 dB	

Parameter/Range	Frequency	CMC ² (±)	Comments
RF Absolute Power – Generate ^{3,6}			
Into 50 Ω (10 to 3) V p-p	0.001 Hz to 100 kHz SWR 1.2:1	0.12 dB	HP 3325A/B with BNC(f)
2.99 V to 1 mV p-p	0.001 Hz to 100 kHz SWR 1.2:1	0.23 dB	
Into 50 Ω (10 to 3) V p-p	100 kHz to 20 MHz SWR ≤ 1.2:1	0.47 dB	HP 3325A/B with BNC(f)
2.999 V to 1 mV p-p	100 kHz to 10 MHz SWR ≤ 1.2:1	0.70 dB	
(2.999 to 0.1) V p-p	(10 to 20) MHz SWR ≤ 1.2:1	0.70 dB	
(99.99 to 1) mV p-p	(10 to 20) MHz SWR ≤ 1.2:1	1.0 dB	
Full Amplitude, 50Ω 13.01 dBm	1 kHz to 25 MHz 200 Hz to 80 MHz	0.14 dB 0.23 dB	HP 3325A/B with BNC(f)
Full Amplitude, 75Ω 11.25 dBm	1 kHz to 25 MHz 200 Hz to 25 MHz	0.14 dB 0.23 dB	
In 2 dB steps, 50 Ω (0 to -18) dBm (-20 to -58) dBm (-60 to -98) dBm	200 Hz to 80 MHz 200 Hz to 80 MHz 200 Hz to 80 MHz	0.28 dB 0.34 dB 0.47 dB	HP 3325A/B with BNC(f)
In 2 dB steps, 75 Ω (0 to -18) dBm	200 Hz to 25 MHz (25 to 80) MHz	0.28 dB 0.41 dB	
(-20 to -58) dBm	200 Hz to 25 MHz (25 to 80) MHz	0.34 dB 0.52 dB	
(-60 to -98) dBm	200 Hz to 25 MHz (25 to 80) MHz	0.47 dB 0.81 dB	
In 0.01 dB steps (0 to -1.99) dBm	200Hz to 80MHz SWR ≤ 1.5:1	0.036 dB	HP 3335A with BNC(f)

Parameter/Range	Frequency	CMC ² (±)	Comments
RF Absolute Power – Generate ^{3,6} – (cont)			
Into 50 Ω (+16 to -119.9) dBm	100 kHz to 2.56 GHz SWR ≤ 1.5:1	1.2 dB	HP 8663A, Type-N(f)
(-120 to -129.9) dBm	100 kHz to 2.56 GHz SWR ≤ 1.5:1	3.5 dB	HP 8663A, Type-N(f)
> +10 dBm	10 MHz to 2 GHz SWR ≤ 1.6:1	1.4 dB	HP 83650B, 2.4 mm(m)
	(≥ 2 to ≤ 20) GHz SWR ≤ 1.6:1	1.5 dB	
> -10 dBm	10 MHz to 2 GHz SWR ≤ 1.6:1	0.72 dB	HP 83650B, 2.4 mm(m)
	(≥ 2 to ≤ 20) GHz SWR ≤ 1.6:1	0.84 dB	
	(> 20 to ≤ 40) GHz SWR ≤ 1.8:1	1.1 dB	
	(> 40 to ≤ 50) GHz SWR ≤ 2:1	1.0 dB	
> -60 dBm	10 MHz to 2 GHz SWR ≤ 1.6:1	1.1 dB	HP 83650B, 2.4 mm(m)
	(≥ 2 to ≤ 20) GHz SWR ≤ 1.6:1	1.2 dB	
	(> 20 to ≤ 40) GHz SWR ≤ 1.8:1	1.4 dB	
	(> 40 to ≤ 50) GHz SWR ≤ 2:1	2.4 dB	
≤ -60 dBm	10 MHz to 2 GHz SWR ≤ 1.6:1	1.7 dB	HP 83650B, 2.4 mm(m)
	(≥ 2 to ≤ 20) GHz SWR ≤ 1.6:1	1.8 dB	
	(> 20 to ≤ 40) GHz SWR ≤ 1.8:1	2.0 dB	
	(> 40 to ≤ 50) GHz SWR ≤ 2:1	3.0 dB	

Parameter/Range	Frequency	CMC ² (±)	Comments
Attenuation – Generate			
Coaxial, 1 dB Step (0 to 11) dB		± dB	
0 dB	50 MHz to 2 GHz (2 to 4) GHz	0.027 0.029	HP 8494G w/ Type-N(f)
1 dB	50 MHz to 2 GHz (2 to 4) GHz	0.027 0.029	
2 dB	50 MHz to 2 GHz (2 to 4) GHz	0.027 0.029	
3 dB	50 MHz to 2 GHz (2 to 4) GHz	0.028 0.029	
4 dB	50 MHz to 2 GHz (2 to 4) GHz	0.028 0.030	
5 dB	50 MHz to 2 GHz (2 to 4) GHz	0.029 0.030	
6 dB	50 MHz to 2 GHz (2 to 4) GHz	0.030 0.030	
7 dB	50 MHz to 2 GHz (2 to 4) GHz	0.030 0.030	
8 dB	50 MHz to 2 GHz (2 to 4) GHz	0.030 0.030	
9 dB	50 MHz to 2 GHz (2 to 4) GHz	0.031 0.030	
10 dB	50 MHz to 2 GHz (2 to 4) GHz	0.031 0.030	
11 dB	50 MHz to 2 GHz (2 to 4) GHz	0.031 0.030	

Parameter/Range	Frequency	CMC ² (±)	Comments
Attenuation – Generate (cont.)		± dB	
Coaxial, 10 dB Step ³			
0 dB	50 MHz to 2 GHz (2 to 4) GHz	0.027 0.029	HP 8496G w/ Type- N(f)
10 dB	50 MHz to 2 GHz (2 to 4) GHz	0.031 0.030	
20 dB	50 MHz to 2 GHz (2 to 4) GHz	0.033 0.032	
30 dB	50 MHz to 2 GHz (2 to 4) GHz	0.036 0.035	
40 dB	50 MHz to 2 GHz (2 to 4) GHz	0.054 0.048	
50 dB	50 MHz to 2 GHz (2 to 4) GHz	0.062 0.057	
60 dB	50 MHz to 2 GHz (2 to 4) GHz	0.062 0.057	
70 dB	50 MHz to 2 GHz (2 to 4) GHz	0.070 0.060	
80 dB	50 MHz to 2 GHz (2 to 4) GHz	0.084 0.068	
90 dB	50 MHz to 2 GHz (2 to 4) GHz	0.091 0.076	
100 dB	50 MHz to 2 GHz (2 to 4) GHz	0.093 0.076	
110 dB	50 MHz to 2 GHz (2 to 4) GHz	0.11 0.077	

Parameter/Range	Frequency	CMC ² (±)	Comments
Attenuation – Generate (cont.)		± dB	
Coaxial, Fixed ³			
3 dB	DC to 2 GHz, SWR < 1.25:1	0.028	HP 8491A/B Type-N
	(2 to 4) GHz, SWR < 1.2:1	0.029	
	(4 to 18) GHz, SWR < 1.2:1	0.053	
6 dB	DC to 2 GHz, SWR < 1.25:1	0.029	
	(2 to 4) GHz, SWR < 1.2:1	0.030	
	(4 to 18) GHz, SWR < 1.2:1	0.053	
10 dB	DC to 2 GHz, SWR < 1.25:1	0.031	
	(2 to 4) GHz, SWR < 1.2:1	0.030	
	(4 to 18) GHz, SWR < 1.2:1	0.053	
20 dB	DC to 2 GHz, SWR < 1.5:1	0.032	
	(2 to 4) GHz, SWR < 1.5:1	0.032	
	(4 to 18) GHz, SWR < 1.5:1	0.053	

Parameter/Range	CMC ² (±)	Comments
Reflection S ₁₁ / S ₂₂ – Measure ³		
30 kHz to 1.3 GHz (0 to 1.0) lin	(± 0.0044 to ±0.022) lin (± 180 to ±1.3) deg	Network analyzer HP8753ES Type-N precision cal kit 85032B
300 kHz to 1.3 GHz (0 to 1.0) lin	(± 0.0071 to ±0.035) lin (± 180 to ±2.0) deg	APC 7mm precision cal kit HP85031B
300 kHz to 1.3 GHz (0 to 1.0) lin	(± 0.0038 to ±0.017) lin (± 180 to ±0.96) deg	Network analyzer HP8753ES Type N precision cal kit 85032B
1.3 GHz to 3 GHz (0 to 1.0) lin	(± 0.0051 to ±0.026) lin (± 180 to ±1.504) deg	
3 GHz to 6 GHz (0 to 1.0) lin	(±0.011 to ±0.050) lin (± 180 to ±3.0) deg	
300 kHz to 1.3 GHz (0 to 1.0) lin	(± 0.0019 to ±0.0072) lin (± 180 to ±0.42) deg	Network analyzer HP8753ES APC 7mm precision cal kit HP85031B
1.3 GHz to 3 GHz (0 to 1.0) lin	(± 0.0036 to ±0.011) lin (± 180 to ±0.59) deg	
3 GHz to 6 GHz (0 to 1.0) lin	(± 0.0057 to ±0.018) lin (± 180 to ±1.1) deg	
45 MHz to 2 GHz (0 to 1) lin	(± 0.0080 to ± 0.0118) lin	Network Analyzer: HP8510B/C
(0 to 0.01) lin	(± 180 to ± 54) deg	S Parameter: HP8517A
(0.01 to 0.1) lin	(± 54 to ± 4.8) deg	Source: HP83650A/B
(0.1 to 0.5) lin	(± 4.8 to ± 1.1) deg	Cable: HP85133E
(0.5 to 1.0) lin	(± 1.1 to ± 0.81) deg	2.4mm Precision Cal Kit HP85056A
(2 to 20) GHz (0 to 1) lin	(± 0.0080 to ± 0.016) lin	
(0 to 0.01) lin	(± 180 to ± 54) deg	
(0.01 to 0.1) lin	(± 54 to ± 5.1) deg	
(0.1 to 0.5) lin	(± 5.1 to ± 1.4) deg	
(0.5 to 1.0) lin	(± 1.4 to ± 1.3) deg	
(20 to 40) GHz (0 to 1) lin	(± 0.013 to ± 0.026) lin	
(0 to 0.01) lin	(± 180 to ± 180) deg	
(0.01 to 0.1) lin	(± 180 to ± 8.0) deg	
(0.1 to 0.5) lin	(± 8.0 to ± 2.3) deg	
(0.5 to 1.0) lin	(± 2.3 to ± 2.2) deg	

Parameter/Range	CMC ² (\pm)	Comments
Reflection S ₁₁ / S ₂₂ – Measure (cont) ³ (40 to 50) GHz (0 to 1) lin (0 to 0.01) lin (0.01 to 0.1) lin (0.1 to 0.5) lin (0.5 to 1.0) lin	(± 0.017 to ± 0.033) lin (± 180 to ± 180) deg (± 180 to ± 11) deg (± 11 to ± 2.9) deg (± 2.9 to ± 2.7) deg	Network Analyzer: HP8510B/C S Parameter: HP8517A Source: HP83650A/B Cable: HP85133E 2.4mm Precision Cal Kit HP85056A
Transmission S ₁₂ / S ₂₁ – Measure ³ 30 kHz to 1.3 GHz (0 to 20) dB (20 to 40) dB (40 to 60) dB 1.3 GHz to 3 GHz (0 to 20) dB (20 to 40) dB (40 to 60) dB 3 GHz to 6 GHz (0 to 20) dB (20 to 40) dB (40 to 60) dB	(± 0.041 to ± 0.063) dB (± 0.42 to ± 0.46) deg (± 0.063 to ± 0.088) dB (± 0.46 to ± 0.62) deg (± 0.088 to ± 0.29) dB (± 0.62 to ± 2.0) deg (± 0.056 to ± 0.076) dB (± 0.79 to ± 1.3) deg (± 0.076 to ± 0.11) dB (± 1.3 to ± 1.4) deg (± 0.11 to ± 0.32) dB (± 1.4 to ± 2.2) deg (± 0.094 to ± 0.13) dB (± 1.2 to ± 1.6) deg (± 0.13 to ± 0.16) dB (± 1.6 to ± 1.7) deg (± 0.16 to ± 0.51) dB (± 1.7 to ± 3.6) deg	Network analyzer 8753ES Type-N precision cal kit 85032B

Parameter/Range	CMC ² (±)	Comments
Transmission S ₁₂ / S ₂₁ – Measure (cont)		
300 kHz to 1.3 GHz (0 to 20) dB	(± 0.028 to ± 0.049) dB (± 0.30 to ± 0.41) deg	Network Analyzer HP8753ES APC 7mm Precision Cal Kit HP85031B
(20 to 40) dB	(± 0.049 to ± 0.075) dB (± 0.41 to ± 0.56) deg	
(40 to 60) dB	(± 0.075 to ± 0.28) dB (± 0.56 to ± 1.9) deg	
1.3 GHz to 3 GHz (0 to 20) dB	(± 0.033 to ± 0.055) dB (± 0.37 to ± 0.47) deg	
(20 to 40) dB	(± 0.055 to ± 0.081) dB (± 0.47 to ± 0.60) deg	
(40 to 60) dB	(± 0.081 to ± 0.32) dB (± 0.60 to ± 2.2) deg	
3 GHz to 6 GHz (0 to 20) dB	(± 0.047 to ± 0.069) dB (± 1.4 to ± 1.2) deg	
(20 to 40) dB	(± 0.069 to ± 0.11) dB (± 1.2 to ± 1.3) deg	
(40 to 60) dB	(± 0.11 to ± 0.49) dB (± 1.3 to ± 3.4) deg	
45 MHz to 2 GHz (0 to 20) dB	(± 0.022 to ± 0.027) dB (± 0.24 to ± 0.29) deg	Network Analyzer HP8510B/C S Parameter HP8517A Source HP83650A/B Cable HP 85133E 2.4mm Precision Cal Kit HP85056A
(20 to 40) dB	(± 0.027 to ± 0.13) dB (± 0.29 to ± 0.92) deg	
(40 to 60) dB	(± 0.13 to ± 1.3) dB (± 0.92 to ± 7.8) deg	
(2 to 20) GHz (0 to 20) dB	(± 0.049 to ± 0.050) dB (± 2.3 to ± 2.4) deg	
(20 to 40) dB	(± 0.050 to ± 0.057) dB (± 2.4 to ± 2.4) deg	
(40 to 60) dB	(± 0.057 to ± 0.25) dB (± 2.4 to ± 3.6) deg	

Parameter/Range	CMC ² (±)	Comments
Transmission S12/ S21 – Measure (cont) ³		
(40 to 50) GHz (0 to 20) dB	(± 0.15 to ± 0.17) dB (± 5.3 to ± 5.9) deg	Network Analyzer HP8510B/C S Parameter HP8517A Source HP83650A/B Cable HP 85133E 2.4mm Precision Cal Kit HP85056A
(20 to 40) dB	(± 0.17 to ± 0.80) dB (± 5.9 to ± 9.8) deg	
(40 to 60) dB	(± 0.80 to ± 17) dB (± 9.8 to ± 64) deg	

III. Time and Frequency

Parameter/Range	Frequency	CMC ^{2, 11} (±)	Comments
Pulse – Generate ³			
Transition Time	10 % to 90 %	130 ps	HP 8133A
Width	150 ps to 10 ns	130 ps	HP 8133A
	10 ns to < 100 ns	(0.013 * Width) + 1.2 ns	HP 8161A
	≥ 100 μs to < 10 ms	(0.013 * Width) + 140 ns	HP 8161A
	≥ 10 ms to < 100 ms	(0.012 * Width) + 20 ns	HP 8161A
RMS Jitter – Period, Delay and Width	≥ 100 ms to < 0.99 ms	(0.012 * Width) + 190 ns	HP 8161A
	2 σ	10 ps	HP 8133A
	Time Interval – Measure ³		
	60 ps to 1 μs	12 ps	HP 54124T, includes transition time; Δt is the time interval in seconds.
	60 ps to 700 ps	13 ps	
	>700 ps to 1.4 ns	1.4 x 10 ⁻³ TI + 13 ps	
	>1.4 ns to 4.4 ns	1.1 x 10 ⁻³ TI + 17 ps	
	>4.4 ns to 9.0 ns	9.5 x 10 ⁻⁴ TI + 26 ps	
	>9.0 ns to 18 ns	8.0 x 10 ⁻⁴ TI + 56 ps	
	>18 ns to 44 ns	7.6 x 10 ⁻⁴ TI + 110 ps	
	>44 ns to 88 ns	7.2 x 10 ⁻⁴ TI + 210 ps	
	>88 ns to 180 ns	6.8 x 10 ⁻⁴ TI + 540 ps	
	>180 ns to 440 ns	7.0 x 10 ⁻⁴ TI + 1000 ps	
	>440 ns to 880 ns	5.4 x 10 ⁻⁴ TI + 2.3 ns	
	>880 ns to 1.0 μs		

Parameter/Equipment	Range	CMC ² (±)	Comments
Time Interval – Measure (cont) ³	10 ns to 10 s	2.9 ns	HP 5334B/HP 5071A cesium beam standard
Frequency – Measuring Equipment ³	10 MHz	10 pHz/Hz	Stanford Research FS700 with option 1 (Loran-C receiver), Datum 8040
Frequency – Measure ³	1 Hz to 40 GHz	50 pHz/Hz	HP 53132A HP 5352B

¹ This laboratory offers commercial calibration service and field calibration service.

² Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ Field calibration service is available for this calibration and this laboratory meets A2LA R104 – *General Requirements: Accreditation of Field Testing and Field Calibration Laboratories* for these calibrations. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

⁴ Based on using the standard at the temperature the Fluke 5720A/5725A was calibrated ($\text{tcal} \pm 5 \text{ }^\circ\text{C}$) and assuming the instrument is zeroed at least every seven days or when the ambient temperature changes more than $5 \text{ }^\circ\text{C}$. For resistance a zero calibration is performed at least every 12 hours within $\pm 1 \text{ }^\circ\text{C}$ of use. Best measurement uncertainty is based upon 1-year specifications and is read as a portion or percent output plus floor specification. The floor specification is expressed as a fixed value or a portion or percent of the range.

⁵ Based on using the HP 3458A at the temperature (tcal) it was calibrated $\pm 5 \text{ }^\circ\text{C}$ and an auto-calibration (ACAL) was performed within the previous 24 hours ($\pm 1 \text{ }^\circ\text{C}$ of ambient temperature). Best measurement uncertainty is based upon 1-year specifications and is read as a portion or percent output plus floor specification. The floor specification is expressed as a fixed value or a portion or percent of the range.

⁶ CMC does not include the mismatch.

⁷ Ranges are based upon the system combination used:

Instrument/System	Ranges
HP 8902A	2.5 MHz to 1.3 GHz Range 1 & 2 – SWR 1.18:1 Range 3 – SWR 1.4:1
HP 8902A w/ HP 11722A	2.5 MHz to 1.3 GHz Range 1 & 2 – SWR 1.33:1 Range 3 – SWR 1.5:1
HP 8902A w/ HP 11792A or HP 11793A	(50 to 1300) MHz SWR 1.15:1 (1.3 to 18) GHz SWR 1.25:1 (18 to 26.5) GHz SWR 1.4:1

⁸ CMCs are based upon the AM depths. For depths between 0 % and 9.99 %, the digit uncertainty (resolution) is 0.01 %. For depths between 10 % and 99.9 % FS, the digit uncertainty (resolution) is 0.1 %.

⁹ CMCs are based upon the peak phase deviations. For deviations between 0 and 3.999, the digit uncertainty (resolution) is 1 Hz. For deviations between 0 and 39.99, the digit uncertainty (resolution) is 10 Hz. For deviations between 40 and 400, the digit uncertainty (resolution) is 100 Hz.

¹⁰ Standards Laboratory parameters and uncertainties are not available for calibrations in the field.

¹¹ *AM* is the amplitude modulation; *FM* is the frequency modulation; and *TI* is the time interval.

¹² All percentages are “of reading” unless otherwise indicated.



The American Association for Laboratory Accreditation

World Class Accreditation

Accredited Laboratory
A2LA has accredited

**AGILENT TECHNOLOGIES CANADA
SERVICE CENTER**

Saint-Laurent, CANADA

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 *General Requirements for the Competence of Testing and Calibration Laboratories*. This laboratory also meets the requirements of ANSI/NCSL Z540-1-1994, ANSI/NCSL Z540.3-2006 and any additional program requirements in the field of calibration. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (*refer to joint ISO-ILAC-IAF Communiqué dated 8 January 2009*).

Presented this 15th day of December 2009.





Peter Mlynar

President & CEO
For the Accreditation Council
Certificate Number 2645.01
Valid to January 31, 2012
Revised August 3, 2010

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.