



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005  
& ANSI/NCSL Z540-1-1994

OMNI-TECH CORPORATION  
 1460 Torrey Road Suite A  
 Fenton, MI 48430-3323  
 Daryl Barlow Phone: 810 750 6474

CALIBRATION

Valid To: February 28, 2018

Certificate Number: 1791.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations<sup>1</sup>:

I. Dimensional

Parameter/Equipment	Range	CMC <sup>2,4</sup> (±)	Comments
Measuring Machines and Machine Tools <sup>3</sup> –			
Linear Displacement Accuracy	Displacement from (0.1 to 10) m	$(0.10 + 1.3L) \mu\text{m}$	Laser
Temperature Compensated Machine	Displacement from (0 to 40) in	$(10 + 2.0L) \mu\text{in}$	Step gage
Non-Temperature Compensated Machine	Displacement from (0 to 40) in	$(10 + 3.1L) \mu\text{in}$	
Measurement Repeatability	Sphere Ø from (12 to 55) mm	0.31 $\mu\text{m}$	Sphere
Volumetric Performance	Ball bar lengths from (0.1 to 1) m	0.36 $\mu\text{m}$	Ball bar

---

<sup>1</sup> This laboratory offers commercial calibration service and field calibration service.

<sup>2</sup> Calibration and Measurement Capability Uncertainties (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

<sup>3</sup> Field calibration service is available for this calibration and this laboratory meets A2LA R104 – *General Requirements: Accreditation of Field Testing and Field Calibration Laboratories* for these calibrations. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

<sup>4</sup> In the statement of CMC,  $L$  is the numerical value of the nominal length of the device measured in inches or meters.



## *Accredited Laboratory*

A2LA has accredited

### **OMNI-TECH CORPORATION**

*Fenton, MI*

for technical competence in the field of

## Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements for ANSI/NCSL Z540-1-1994 and any additional program requirements in the field of calibration. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (*refer to joint ISO-ILAC-IAF Communiqué dated 8 January 2009*).



Presented this 20<sup>th</sup> day of November 2015.

A handwritten signature in black ink, reading "Peter Abney".

President & CEO  
For the Accreditation Council  
Certificate Number 1791.01  
Valid to February 28, 2018

*For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.*